

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120825\
 Data File : PO115642.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Dec 2025 14:33
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 09 00:47:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 05 02:11:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.637	378.8E6	234.9E6	48.771	48.202
2) SA Decachlor...	9.910	8.620	262.3E6	202.6E6	49.848	47.124
Target Compounds						
21) L5 AR-1248-1	5.474	4.713	74947036	56609538	493.830	506.215
22) L5 AR-1248-2	5.744	4.948	106.4E6	75703290	530.076	497.101
23) L5 AR-1248-3	5.944	4.990	118.4E6	80139183	480.143	499.226
24) L5 AR-1248-4	6.341	5.160	143.4E6	93338707	486.259	498.748
25) L5 AR-1248-5	6.379	5.550	137.8E6	97345621	532.089	491.756



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120825\
Data File : PO115642.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Dec 2025 14:33
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 09 00:47:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 05 02:11:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: -0.001 min
Response: 378817158
Conc: 48.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.001 min
Response: 234903250
Conc: 48.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.910 min
Delta R.T.: -0.002 min
Response: 262254766
Conc: 49.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.620 min
Delta R.T.: -0.002 min
Response: 202644593
Conc: 47.12 ng/ml



#21 AR-1248-1


R.T.: 5.474 min
Delta R.T.: 0.000 min
Response: 74947036
Conc: 493.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.713 min
Delta R.T.: 0.000 min
Response: 56609538
Conc: 506.22 ng/ml




#22 AR-1248-2

R.T.: 5.744 min
Delta R.T.: 0.000 min
Response: 106447189
Conc: 530.08 ng/ml



#22 AR-1248-2


R.T.: 4.948 min
Delta R.T.: -0.001 min
Response: 75703290
Conc: 497.10 ng/ml



#23 AR-1248-3


R.T.: 5.944 min
Delta R.T.: 0.000 min
Response: 118402884
Conc: 480.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.990 min
Delta R.T.: -0.001 min
Response: 80139183
Conc: 499.23 ng/ml




#24 AR-1248-4

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 143375845
Conc: 486.26 ng/ml



#24 AR-1248-4


R.T.: 5.160 min
Delta R.T.: 0.000 min
Response: 93338707
Conc: 498.75 ng/ml



#25 AR-1248-5

R.T.: 6.379 min
Delta R.T.: 0.000 min
Response: 137771603
Conc: 532.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.550 min
Delta R.T.: 0.000 min
Response: 97345621
Conc: 491.76 ng/ml