

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120922\
 Data File : PO091227.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Dec 2022 17:42
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 00:04:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.297	3.474	157.5E6	52641995	46.573	50.156
2) SA Decachlor...	10.043	8.476	115.4E6	47314419	43.435	48.734
Target Compounds						
26) L6 AR-1254-1	6.334	5.363	54953853	27556790	437.561	470.198
27) L6 AR-1254-2	6.555	5.512	83007697	25091040	440.225	470.132
28) L6 AR-1254-3	6.923	5.916	84071309	38391059	448.530	471.220
29) L6 AR-1254-4	7.210	6.146	62765685	23450027	454.363	479.283
30) L6 AR-1254-5	7.632	6.566	68576765	35259149	446.962	472.994



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120922\
Data File : PO091227.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Dec 2022 17:42
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 00:04:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 157468238
Conc: 46.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.474 min
Delta R.T.: -0.002 min
Response: 52641995
Conc: 50.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.043 min
Delta R.T.: -0.007 min
Response: 115370939
Conc: 43.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.476 min
Delta R.T.: -0.006 min
Response: 47314419
Conc: 48.73 ng/ml



#26 AR-1254-1


R.T.: 6.334 min
Delta R.T.: -0.004 min
Response: 54953853
Conc: 437.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.363 min
Delta R.T.: -0.005 min
Response: 27556790
Conc: 470.20 ng/ml




#27 AR-1254-2

R.T.: 6.555 min
Delta R.T.: -0.004 min
Response: 83007697
Conc: 440.22 ng/ml



#27 AR-1254-2


R.T.: 5.512 min
Delta R.T.: -0.005 min
Response: 25091040
Conc: 470.13 ng/ml



#28 AR-1254-3


R.T.: 6.923 min
Delta R.T.: -0.005 min
Response: 84071309
Conc: 448.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.916 min
Delta R.T.: -0.006 min
Response: 38391059
Conc: 471.22 ng/ml




#29 AR-1254-4

R.T.: 7.210 min
Delta R.T.: -0.006 min
Response: 62765685
Conc: 454.36 ng/ml



#29 AR-1254-4


R.T.: 6.146 min
Delta R.T.: -0.006 min
Response: 23450027
Conc: 479.28 ng/ml



#30 AR-1254-5

R.T.: 7.632 min
Delta R.T.: -0.006 min
Response: 68576765
Conc: 446.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.566 min
Delta R.T.: -0.006 min
Response: 35259149
Conc: 472.99 ng/ml