

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121020\
 Data File : PO073646.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Dec 2020 1:36
 Operator : DD\AJ
 Sample : AR1262ICV500
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO121020AR1262

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 04:03:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 11 03:58:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.941	4.004	5042561	2987237	55.154	55.694
2)	SA Decachlor...	10.968	9.270	4793752	2845297	58.008	60.809
Target Compounds							
3)	L1 AR-1016-1	6.266	5.253	53079	34511	15.106	15.799
4)	L1 AR-1016-2	6.289	5.272	76354	51122	15.622	16.894
5)	L1 AR-1016-3	6.354	5.462	46295	24549	15.716	15.058
6)	L1 AR-1016-4	6.463	5.513	37215	34562	15.100	26.209 #
7)	L1 AR-1016-5	6.778	5.740	35874	26249	15.109	16.019
9)	L2 AR-1221-2	5.294	4.354	68571	41857	87.296	81.406
10)	L2 AR-1221-3	5.379	0.000	28866	0	11.866	N.D. #
11)	L3 AR-1232-1	5.379	0.000	28866	0	13.999	N.D. #
12)	L3 AR-1232-2	5.969	5.272	37335	51122	34.918	38.470
13)	L3 AR-1232-3	6.289	5.462	76354	24549	35.580	33.733
14)	L3 AR-1232-4	6.463	5.558	37215	21096	34.944	34.563
15)	L3 AR-1232-5	6.560	5.740	51139	26249	71.425	39.276 #
16)	L4 AR-1242-1	6.266	5.253	53079	34511	18.728	19.622
17)	L4 AR-1242-2	6.289	5.272	76354	51122	19.604	20.902
18)	L4 AR-1242-3	6.354	5.462	46295	24549	19.517	18.758
19)	L4 AR-1242-4	6.463	5.558	37215	21096	18.880	16.857
20)	L4 AR-1242-5	7.249	6.119	48500	243595	22.209	142.905 #
21)	L5 AR-1248-1	6.266	5.253	53079	34511	24.428	25.206
22)	L5 AR-1248-2	6.560	5.513	51139	34562	18.300	18.628
23)	L5 AR-1248-3	6.778	5.558	35874	21096	10.954	11.244
24)	L5 AR-1248-4	7.209	5.740	51830	26249	12.960	11.662
25)	L5 AR-1248-5	7.249	6.160	48500	30126	12.825	12.966
26)	L6 AR-1254-1	7.178	6.119	313642	243595	82.629	68.928
27)	L6 AR-1254-2	7.408	6.278	394076	262330	66.840	84.026 #
28)	L6 AR-1254-3	7.792	6.714f	161728	1006673	26.395	205.611 #
29)	L6 AR-1254-4	8.086	6.932	142463	68402	27.157	19.504 #
30)	L6 AR-1254-5	8.515	7.358	2212965	1226412	429.658	278.122 #
31)	L7 AR-1260-1	7.959	6.831	1866277	1342159	441.573	434.173
32)	L7 AR-1260-2	8.225	7.028	2453488	1760080	436.734	433.890
33)	L7 AR-1260-3	8.591	7.183	3690579	1607063	807.367	460.435 #
34)	L7 AR-1260-4	8.824	7.662	3020378	2080981	637.962	689.125
35)	L7 AR-1260-5	9.149	7.908	6377802	4403989	639.074	589.552
36)	L8 AR-1262-1	8.591	7.358	3690579	1226412	492.985	490.604
37)	L8 AR-1262-2	9.149	7.908	6377802	4403989	497.259	500.660
38)	L8 AR-1262-3	9.460	8.193	4201849	1764418	495.277	487.963
39)	L8 AR-1262-4	9.553	8.256	1768953	3025559	528.905	491.715
40)	L8 AR-1262-5	10.220	8.750	2283046	1424408	505.099	500.325
41)	L9 AR-1268-1	9.460	8.193	4201849	1764418	295.009	183.344 #
42)	L9 AR-1268-2	9.553	8.256	1768953	3025559	143.502	367.242 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121020\
 Data File : PO073646.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Dec 2020 1:36
 Operator : DD\AJ
 Sample : AR1262ICV500
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO121020AR1262

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 04:03:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 11 03:58:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	9.782	8.464	196755	128378	18.129	18.182
44) L9	AR-1268-4	10.220	8.750	2283046	1424408	487.360	481.217
45) L9	AR-1268-5	10.633	9.028	824642	523210	25.198	26.043



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121020\
Data File : PO073646.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Dec 2020 1:36
Operator : DD\AJ
Sample : AR1262ICV500
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 04:03:02 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 11 03:58:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.941 min
Delta R.T.: 0.000 min
Response: 5042561
Conc: 55.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#1 Tetrachloro-m-xylene

R.T.: 4.004 min
Delta R.T.: 0.000 min
Response: 2987237
Conc: 55.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.968 min
Delta R.T.: 0.000 min
Response: 4793752
Conc: 58.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.270 min
Delta R.T.: -0.002 min
Response: 2845297
Conc: 60.81 ng/ml



#3 AR-1016-1


R.T.: 6.266 min
Delta R.T.: 0.000 min
Response: 53079
Conc: 15.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#3 AR-1016-1

R.T.: 5.253 min
Delta R.T.: 0.000 min
Response: 34511
Conc: 15.80 ng/ml




#4 AR-1016-2

R.T.: 6.289 min
Delta R.T.: 0.000 min
Response: 76354
Conc: 15.62 ng/ml



#4 AR-1016-2


R.T.: 5.272 min
Delta R.T.: 0.000 min
Response: 51122
Conc: 16.89 ng/ml



#5 AR-1016-3


R.T.: 6.354 min
Delta R.T.: -0.001 min
Response: 46295
Conc: 15.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#5 AR-1016-3

R.T.: 5.462 min
Delta R.T.: 0.000 min
Response: 24549
Conc: 15.06 ng/ml




#6 AR-1016-4

R.T.: 6.463 min
Delta R.T.: 0.000 min
Response: 37215
Conc: 15.10 ng/ml



#6 AR-1016-4


R.T.: 5.513 min
Delta R.T.: -0.001 min
Response: 34562
Conc: 26.21 ng/ml



#7 AR-1016-5


R.T.: 6.778 min
Delta R.T.: 0.000 min
Response: 35874
Conc: 15.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#7 AR-1016-5

R.T.: 5.740 min
Delta R.T.: -0.002 min
Response: 26249
Conc: 16.02 ng/ml




#9 AR-1221-2

R.T.: 5.294 min
Delta R.T.: 0.004 min
Response: 68571
Conc: 87.30 ng/ml



#9 AR-1221-2


R.T.: 4.354 min
Delta R.T.: -0.004 min
Response: 41857
Conc: 81.41 ng/ml



#10 AR-1221-3


R.T.: 5.379 min
Delta R.T.: 0.001 min
Response: 28866
Conc: 11.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.445 min
Response: 0
Conc: N.D.




#11 AR-1232-1

R.T.: 5.379 min
Delta R.T.: 0.002 min
Response: 28866
Conc: 14.00 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 4.445 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.969 min
Delta R.T.: 0.001 min
Response: 37335
Conc: 34.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#12 AR-1232-2

R.T.: 5.272 min
Delta R.T.: 0.000 min
Response: 51122
Conc: 38.47 ng/ml




#13 AR-1232-3

R.T.: 6.289 min
Delta R.T.: 0.001 min
Response: 76354
Conc: 35.58 ng/ml



#13 AR-1232-3


R.T.: 5.462 min
Delta R.T.: 0.000 min
Response: 24549
Conc: 33.73 ng/ml



#14 AR-1232-4


R.T.: 6.463 min
Delta R.T.: 0.001 min
Response: 37215
Conc: 34.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#14 AR-1232-4

R.T.: 5.558 min
Delta R.T.: 0.000 min
Response: 21096
Conc: 34.56 ng/ml




#15 AR-1232-5

R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 51139
Conc: 71.42 ng/ml



#15 AR-1232-5


R.T.: 5.740 min
Delta R.T.: -0.001 min
Response: 26249
Conc: 39.28 ng/ml



#16 AR-1242-1


R.T.: 6.266 min
Delta R.T.: 0.000 min
Response: 53079
Conc: 18.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#16 AR-1242-1

R.T.: 5.253 min
Delta R.T.: -0.002 min
Response: 34511
Conc: 19.62 ng/ml




#17 AR-1242-2

R.T.: 6.289 min
Delta R.T.: -0.001 min
Response: 76354
Conc: 19.60 ng/ml



#17 AR-1242-2


R.T.: 5.272 min
Delta R.T.: -0.001 min
Response: 51122
Conc: 20.90 ng/ml



#18 AR-1242-3


R.T.: 6.354 min
Delta R.T.: -0.003 min
Response: 46295
Conc: 19.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#18 AR-1242-3

R.T.: 5.462 min
Delta R.T.: -0.001 min
Response: 24549
Conc: 18.76 ng/ml




#19 AR-1242-4

R.T.: 6.463 min
Delta R.T.: 0.000 min
Response: 37215
Conc: 18.88 ng/ml



#19 AR-1242-4


R.T.: 5.558 min
Delta R.T.: -0.002 min
Response: 21096
Conc: 16.86 ng/ml



#20 AR-1242-5


R.T.: 7.249 min
Delta R.T.: 0.000 min
Response: 48500
Conc: 22.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#20 AR-1242-5

R.T.: 6.119 min
Delta R.T.: -0.001 min
Response: 243595
Conc: 142.91 ng/ml




#21 AR-1248-1

R.T.: 6.266 min
Delta R.T.: 0.002 min
Response: 53079
Conc: 24.43 ng/ml



#21 AR-1248-1


R.T.: 5.253 min
Delta R.T.: 0.000 min
Response: 34511
Conc: 25.21 ng/ml



#22 AR-1248-2


R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 51139
Conc: 18.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#22 AR-1248-2

R.T.: 5.513 min
Delta R.T.: 0.000 min
Response: 34562
Conc: 18.63 ng/ml




#23 AR-1248-3

R.T.: 6.778 min
Delta R.T.: 0.000 min
Response: 35874
Conc: 10.95 ng/ml



#23 AR-1248-3


R.T.: 5.558 min
Delta R.T.: 0.000 min
Response: 21096
Conc: 11.24 ng/ml



#24 AR-1248-4


R.T.: 7.209 min
Delta R.T.: 0.002 min
Response: 51830
Conc: 12.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#24 AR-1248-4

R.T.: 5.740 min
Delta R.T.: -0.001 min
Response: 26249
Conc: 11.66 ng/ml




#25 AR-1248-5

R.T.: 7.249 min
Delta R.T.: 0.002 min
Response: 48500
Conc: 12.82 ng/ml



#25 AR-1248-5


R.T.: 6.160 min
Delta R.T.: -0.001 min
Response: 30126
Conc: 12.97 ng/ml



#26 AR-1254-1


R.T.: 7.178 min
Delta R.T.: 0.000 min
Response: 313642
Conc: 82.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#26 AR-1254-1

R.T.: 6.119 min
Delta R.T.: 0.000 min
Response: 243595
Conc: 68.93 ng/ml




#27 AR-1254-2

R.T.: 7.408 min
Delta R.T.: 0.000 min
Response: 394076
Conc: 66.84 ng/ml



#27 AR-1254-2


R.T.: 6.278 min
Delta R.T.: 0.000 min
Response: 262330
Conc: 84.03 ng/ml



#28 AR-1254-3


R.T.: 7.792 min
Delta R.T.: 0.001 min
Response: 161728
Conc: 26.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#28 AR-1254-3

R.T.: 6.714 min
Delta R.T.: 0.020 min
Response: 1006673
Conc: 205.61 ng/ml




#29 AR-1254-4

R.T.: 8.086 min
Delta R.T.: -0.001 min
Response: 142463
Conc: 27.16 ng/ml



#29 AR-1254-4


R.T.: 6.932 min
Delta R.T.: 0.000 min
Response: 68402
Conc: 19.50 ng/ml



#30 AR-1254-5


R.T.: 8.515 min
Delta R.T.: -0.001 min
Response: 2212965
Conc: 429.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#30 AR-1254-5

R.T.: 7.358 min
Delta R.T.: 0.000 min
Response: 1226412
Conc: 278.12 ng/ml




#31 AR-1260-1

R.T.: 7.959 min
Delta R.T.: 0.000 min
Response: 1866277
Conc: 441.57 ng/ml



#31 AR-1260-1


R.T.: 6.831 min
Delta R.T.: 0.000 min
Response: 1342159
Conc: 434.17 ng/ml



#32 AR-1260-2


R.T.: 8.225 min
Delta R.T.: 0.000 min
Response: 2453488
Conc: 436.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#32 AR-1260-2

R.T.: 7.028 min
Delta R.T.: 0.000 min
Response: 1760080
Conc: 433.89 ng/ml




#33 AR-1260-3

R.T.: 8.591 min
Delta R.T.: 0.000 min
Response: 3690579
Conc: 807.37 ng/ml



#33 AR-1260-3


R.T.: 7.183 min
Delta R.T.: 0.001 min
Response: 1607063
Conc: 460.43 ng/ml



#34 AR-1260-4


R.T.: 8.824 min
Delta R.T.: 0.001 min
Response: 3020378
Conc: 637.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#34 AR-1260-4

R.T.: 7.662 min
Delta R.T.: 0.000 min
Response: 2080981
Conc: 689.13 ng/ml




#35 AR-1260-5

R.T.: 9.149 min
Delta R.T.: 0.000 min
Response: 6377802
Conc: 639.07 ng/ml



#35 AR-1260-5


R.T.: 7.908 min
Delta R.T.: 0.000 min
Response: 4403989
Conc: 589.55 ng/ml



#36 AR-1262-1


R.T.: 8.591 min
Delta R.T.: 0.001 min
Response: 3690579
Conc: 492.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#36 AR-1262-1

R.T.: 7.358 min
Delta R.T.: 0.001 min
Response: 1226412
Conc: 490.60 ng/ml




#37 AR-1262-2

R.T.: 9.149 min
Delta R.T.: 0.000 min
Response: 6377802
Conc: 497.26 ng/ml



#37 AR-1262-2


R.T.: 7.908 min
Delta R.T.: 0.001 min
Response: 4403989
Conc: 500.66 ng/ml



#38 AR-1262-3


R.T.: 9.460 min
Delta R.T.: 0.000 min
Response: 4201849
Conc: 495.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#38 AR-1262-3

R.T.: 8.193 min
Delta R.T.: 0.000 min
Response: 1764418
Conc: 487.96 ng/ml




#39 AR-1262-4

R.T.: 9.553 min
Delta R.T.: 0.000 min
Response: 1768953
Conc: 528.90 ng/ml



#39 AR-1262-4


R.T.: 8.256 min
Delta R.T.: 0.000 min
Response: 3025559
Conc: 491.71 ng/ml



#40 AR-1262-5


R.T.: 10.220 min
Delta R.T.: 0.000 min
Response: 2283046
Conc: 505.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#40 AR-1262-5

R.T.: 8.750 min
Delta R.T.: 0.000 min
Response: 1424408
Conc: 500.32 ng/ml




#41 AR-1268-1

R.T.: 9.460 min
Delta R.T.: 0.000 min
Response: 4201849
Conc: 295.01 ng/ml



#41 AR-1268-1


R.T.: 8.193 min
Delta R.T.: 0.001 min
Response: 1764418
Conc: 183.34 ng/ml



#42 AR-1268-2


R.T.: 9.553 min
Delta R.T.: -0.003 min
Response: 1768953
Conc: 143.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#42 AR-1268-2

R.T.: 8.256 min
Delta R.T.: 0.000 min
Response: 3025559
Conc: 367.24 ng/ml




#43 AR-1268-3

R.T.: 9.782 min
Delta R.T.: 0.000 min
Response: 196755
Conc: 18.13 ng/ml



#43 AR-1268-3


R.T.: 8.464 min
Delta R.T.: 0.000 min
Response: 128378
Conc: 18.18 ng/ml



#44 AR-1268-4


R.T.: 10.220 min
Delta R.T.: 0.000 min
Response: 2283046
Conc: 487.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121020AR1262




#44 AR-1268-4

R.T.: 8.750 min
Delta R.T.: 0.002 min
Response: 1424408
Conc: 481.22 ng/ml



#45 AR-1268-5

R.T.: 10.633 min
Delta R.T.: 0.000 min
Response: 824642
Conc: 25.20 ng/ml



#45 AR-1268-5

R.T.: 9.028 min
Delta R.T.: 0.000 min
Response: 523210
Conc: 26.04 ng/ml