

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO121020\
 Data File : PO073636.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Dec 2020 22:44
 Operator : DD\AJ
 Sample : AR1262ICC050
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1262ICC050

Manual Integrations
 APPROVED

Ankita
 12/11/2020 4:56:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 02:59:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO121020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 11 02:57:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.939	4.002	524000	331601	5.192	5.540
2) SA Decachlor...	10.965	9.268	525201	342195	5.615	6.108
Target Compounds						
36) L8 AR-1262-1	8.588	7.356	424235	146864	57.757	60.232
37) L8 AR-1262-2	9.147	7.906	715086	477677	56.460	54.837
38) L8 AR-1262-3	9.456	8.191	475169	211464	56.586m	60.327
39) L8 AR-1262-4	9.550	8.255	181140	351861	54.480m	58.602
40) L8 AR-1262-5	10.219	8.748	242599	161646	53.963	57.821

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121020\
Data File : PO073636.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Dec 2020 22:44
Operator : DD\AJ
Sample : AR1262ICC050
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
12/11/2020 4:56:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 02:59:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 11 02:57:39 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.939 min
Delta R.T.: 0.000 min
Response: 524000
Conc: 5.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
12/11/2020 4:56:45 PM




#1 Tetrachloro-m-xylene

R.T.: 4.002 min
Delta R.T.: -0.001 min
Response: 331601
Conc: 5.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.965 min
Delta R.T.: 0.001 min
Response: 525201
Conc: 5.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.268 min
Delta R.T.: -0.001 min
Response: 342195
Conc: 6.11 ng/ml




#36 AR-1262-1

R.T.: 8.588 min
Delta R.T.: -0.001 min
Response: 424235
Conc: 57.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
12/11/2020 4:56:45 PM




#36 AR-1262-1

R.T.: 7.356 min
Delta R.T.: 0.000 min
Response: 146864
Conc: 60.23 ng/ml




#37 AR-1262-2

R.T.: 9.147 min
Delta R.T.: 0.000 min
Response: 715086
Conc: 56.46 ng/ml



#37 AR-1262-2

R.T.: 7.906 min
Delta R.T.: 0.000 min
Response: 477677
Conc: 54.84 ng/ml




#38 AR-1262-3

R.T.: 9.456 min
Delta R.T.: -0.002 min
Response: 475169
Conc: 56.59 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
12/11/2020 4:56:45 PM




#38 AR-1262-3

R.T.: 8.191 min
Delta R.T.: 0.000 min
Response: 211464
Conc: 60.33 ng/ml




#39 AR-1262-4

R.T.: 9.550 min
Delta R.T.: -0.002 min
Response: 181140
Conc: 54.48 ng/ml m



#39 AR-1262-4

R.T.: 8.255 min
Delta R.T.: 0.000 min
Response: 351861
Conc: 58.60 ng/ml




#40 AR-1262-5

R.T.: 10.219 min
Delta R.T.: 0.001 min
Response: 242599
Conc: 53.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050

Manual Integrations
APPROVED

Ankita
12/11/2020 4:56:45 PM



#40 AR-1262-5

R.T.: 8.748 min
Delta R.T.: 0.000 min
Response: 161646
Conc: 57.82 ng/ml