

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121021\
 Data File : PO083539.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Dec 2021 11:13
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 01:31:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 07 13:18:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.940	3.945	2082100	650874	21.824	20.925
2) SA Decachlor...	11.018	9.245	1532589	606481	21.454	18.590
Target Compounds						
8) L2 AR-1221-1	5.184	0.000	51358	0	51.071	N.D. #
9) L2 AR-1221-2	5.297	0.000	23915	0	33.203	N.D. #
10) L2 AR-1221-3	0.000	4.387	0	24828	N.D.	26.731 #
11) L3 AR-1232-1	0.000	4.387	0	24828	N.D.	31.409 #
28) L6 AR-1254-3	7.827f	0.000	27428	0	5.498	N.D. #
29) L6 AR-1254-4	8.096	0.000	7534	0	2.114	N.D. #
32) L7 AR-1260-2	8.217f	6.999	11030	25277	2.643	10.343 #
35) L7 AR-1260-5	9.186	0.000	51897	0	7.473	N.D. #
37) L8 AR-1262-2	9.186	0.000	51897	0	7.144	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121021\
Data File : PO083539.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Dec 2021 11:13
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 01:31:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120621.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 07 13:18:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.940 min
Delta R.T.: -0.003 min
Response: 2082100
Conc: 21.82 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 650874
Conc: 20.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.018 min
Delta R.T.: -0.008 min
Response: 1532589
Conc: 21.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.245 min
Delta R.T.: -0.003 min
Response: 606481
Conc: 18.59 ng/ml



#8 AR-1221-1


R.T.: 5.184 min
Delta R.T.: -0.010 min
Response: 51358
Conc: 51.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.202 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.297 min
Delta R.T.: 0.003 min
Response: 23915
Conc: 33.20 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.301 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 5.382 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 4.387 min
Delta R.T.: -0.002 min
Response: 24828
Conc: 26.73 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 5.384 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.387 min
Delta R.T.: -0.003 min
Response: 24828
Conc: 31.41 ng/ml



#28 AR-1254-3


R.T.: 7.827 min
Delta R.T.: 0.021 min
Response: 27428
Conc: 5.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.654 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 8.096 min
Delta R.T.: -0.008 min
Response: 7534
Conc: 2.11 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.895 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.217 min
Delta R.T.: -0.024 min
Response: 11030
Conc: 2.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.999 min
Delta R.T.: 0.009 min
Response: 25277
Conc: 10.34 ng/ml




#35 AR-1260-5

R.T.: 9.186 min
Delta R.T.: 0.014 min
Response: 51897
Conc: 7.47 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.878 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 9.186 min
Delta R.T.: 0.013 min
Response: 51897
Conc: 7.14 ng/ml

Instrument :
ECD_O
ClientSampleId :



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.879 min
Response: 0
Conc: N.D.