

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121021\
 Data File : PO083557.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Dec 2021 19:01
 Operator : AJ\MA
 Sample : M4992-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 01:38:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 07 13:18:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.941	3.946	1826481	590931	19.144	18.997
2) SA Decachlor...	11.017	9.243	1279304	487347	17.908	14.938
Target Compounds						
8) L2 AR-1221-1	5.196	0.000	109846	0	109.233	N.D. #
9) L2 AR-1221-2	5.293	4.272f	22406	24762	31.107	86.016 #
12) L3 AR-1232-2	5.989	0.000	9570	0	9.701	N.D. #
30) L6 AR-1254-5	8.518f	0.000	2835	0	0.790	N.D. #
33) L7 AR-1260-3	8.596	0.000	3294	0	1.051	N.D. #
34) L7 AR-1260-4	8.852	0.000	503	0	0.144	N.D. #
35) L7 AR-1260-5	9.156f	0.000	6085	0	0.876	N.D. #
36) L8 AR-1262-1	8.596	0.000	3294	0	0.793	N.D. #
37) L8 AR-1262-2	9.156f	0.000	6085	0	0.838	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121021\
Data File : PO083557.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Dec 2021 19:01
Operator : AJ\MA
Sample : M4992-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 01:38:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120621.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 07 13:18:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.941 min
Delta R.T.: -0.002 min
Response: 1826481
Conc: 19.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.946 min
Delta R.T.: 0.000 min
Response: 590931
Conc: 19.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.017 min
Delta R.T.: -0.008 min
Response: 1279304
Conc: 17.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.243 min
Delta R.T.: -0.005 min
Response: 487347
Conc: 14.94 ng/ml



#8 AR-1221-1


R.T.: 5.196 min
Delta R.T.: 0.002 min
Response: 109846
Conc: 109.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.202 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 22406
Conc: 31.11 ng/ml



#9 AR-1221-2


R.T.: 4.272 min
Delta R.T.: -0.029 min
Response: 24762
Conc: 86.02 ng/ml



#12 AR-1232-2


R.T.: 5.989 min
Delta R.T.: 0.012 min
Response: 9570
Conc: 9.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.221 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.518 min
Delta R.T.: -0.017 min
Response: 2835
Conc: 0.79 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 7.324 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 8.596 min
Delta R.T.: -0.013 min
Response: 3294
Conc: 1.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 7.144 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.852 min
Delta R.T.: 0.011 min
Response: 503
Conc: 0.14 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 7.628 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 9.156 min
Delta R.T.: -0.016 min
Response: 6085
Conc: 0.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.878 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 8.596 min
Delta R.T.: -0.014 min
Response: 3294
Conc: 0.79 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 7.324 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 9.156 min
Delta R.T.: -0.017 min
Response: 6085
Conc: 0.84 ng/ml

Instrument :
ECD_O
ClientSampleId :



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.879 min
Response: 0
Conc: N.D.