

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121021\
 Data File : PO083568.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Dec 2021 22:46
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 01:43:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 07 13:18:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.940	3.943	1990675	659327	20.865	21.196
2) SA Decachlor...	11.014	9.242	1505995	643064	21.081	19.712
Target Compounds						
8) L2 AR-1221-1	5.183	0.000	50025	0	49.746	N.D. #
9) L2 AR-1221-2	5.295	0.000	26726	0	37.105	N.D. #
10) L2 AR-1221-3	0.000	4.386	0	23257	N.D.	25.039 #
11) L3 AR-1232-1	0.000	4.386	0	23257	N.D.	29.421 #
28) L6 AR-1254-3	7.824f	0.000	19014	0	3.812	N.D. #
30) L6 AR-1254-5	8.536	0.000	18196	0	5.073	N.D. #
32) L7 AR-1260-2	8.214f	0.000	16822	0	4.030	N.D. #
33) L7 AR-1260-3	8.596	0.000	764	0	0.244	N.D. #
34) L7 AR-1260-4	8.836	0.000	13662	0	3.902	N.D. #
35) L7 AR-1260-5	9.174	0.000	2115	0	0.304	N.D. #
36) L8 AR-1262-1	8.596	0.000	764	0	0.184	N.D. #
37) L8 AR-1262-2	9.174	0.000	2115	0	0.291	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121021\
Data File : P0083568.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Dec 2021 22:46
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 01:43:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120621.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 07 13:18:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.940 min
Delta R.T.: -0.003 min
Response: 1990675
Conc: 20.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.943 min
Delta R.T.: -0.002 min
Response: 659327
Conc: 21.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.014 min
Delta R.T.: -0.011 min
Response: 1505995
Conc: 21.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.242 min
Delta R.T.: -0.006 min
Response: 643064
Conc: 19.71 ng/ml



#8 AR-1221-1


R.T.: 5.183 min
Delta R.T.: -0.011 min
Response: 50025
Conc: 49.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.202 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.295 min
Delta R.T.: 0.000 min
Response: 26726
Conc: 37.11 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.301 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 5.382 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 4.386 min
Delta R.T.: -0.003 min
Response: 23257
Conc: 25.04 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T.: 5.384 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.386 min
Delta R.T.: -0.004 min
Response: 23257
Conc: 29.42 ng/ml



#28 AR-1254-3


R.T.: 7.824 min
Delta R.T.: 0.018 min
Response: 19014
Conc: 3.81 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.654 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.536 min
Delta R.T.: 0.000 min
Response: 18196
Conc: 5.07 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 7.324 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.214 min
Delta R.T.: -0.027 min
Response: 16822
Conc: 4.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 6.990 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.596 min
Delta R.T.: -0.012 min
Response: 764
Conc: 0.24 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.144 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 8.836 min
Delta R.T.: -0.005 min
Response: 13662
Conc: 3.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.628 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 9.174 min
Delta R.T.: 0.002 min
Response: 2115
Conc: 0.30 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T. : 7.878 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 8.596 min
Delta R.T.: -0.014 min
Response: 764
Conc: 0.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.324 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 9.174 min
Delta R.T.: 0.000 min
Response: 2115
Conc: 0.29 ng/ml



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.879 min
Response: 0
Conc: N.D.