

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121024\
 Data File : P0108433.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2024 10:38
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 00:29:47 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 05:58:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.708	3.706	468.4E6	271.4E6	53.835	53.537
2) SA Decachlor...	8.782	8.732	335.3E6	199.3E6	45.912	51.421
Target Compounds						
26) L6 AR-1254-1	5.617	5.598	253.5E6	138.6E6	493.477	511.905
27) L6 AR-1254-2	5.766	5.746	222.2E6	121.5E6	487.759	506.359
28) L6 AR-1254-3	6.173	6.150	339.0E6	178.9E6	470.739	476.375
29) L6 AR-1254-4	6.403	6.377	220.0E6	114.9E6	501.266	527.850
30) L6 AR-1254-5	6.823	6.796	299.6E6	168.9E6	471.786	523.329



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121024\
Data File : PO108433.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Dec 2024 10:38
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 00:29:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 05:58:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.708 min
Delta R.T.: -0.002 min
Response: 468371422
Conc: 53.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.706 min
Delta R.T.: -0.002 min
Response: 271399162
Conc: 53.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.782 min
Delta R.T.: -0.008 min
Response: 335344611
Conc: 45.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.732 min
Delta R.T.: -0.009 min
Response: 199251422
Conc: 51.42 ng/ml



#26 AR-1254-1


R.T.: 5.617 min
Delta R.T.: -0.002 min
Response: 253469710
Conc: 493.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.598 min
Delta R.T.: -0.004 min
Response: 138640225
Conc: 511.90 ng/ml




#27 AR-1254-2

R.T.: 5.766 min
Delta R.T.: -0.002 min
Response: 222153877
Conc: 487.76 ng/ml



#27 AR-1254-2


R.T.: 5.746 min
Delta R.T.: -0.004 min
Response: 121536543
Conc: 506.36 ng/ml



#28 AR-1254-3


R.T.: 6.173 min
Delta R.T.: -0.002 min
Response: 338992542
Conc: 470.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.150 min
Delta R.T.: -0.005 min
Response: 178883932
Conc: 476.37 ng/ml




#29 AR-1254-4

R.T.: 6.403 min
Delta R.T.: -0.002 min
Response: 219962007
Conc: 501.27 ng/ml



#29 AR-1254-4


R.T.: 6.377 min
Delta R.T.: -0.004 min
Response: 114900814
Conc: 527.85 ng/ml



#30 AR-1254-5

R.T.: 6.823 min
Delta R.T.: -0.002 min
Response: 299641950
Conc: 471.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.796 min
Delta R.T.: -0.005 min
Response: 168909488
Conc: 523.33 ng/ml