

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121120\
 Data File : PO073705.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Dec 2020 1:24
 Operator : DD\AJ
 Sample : L5033-06MSD
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CS-48MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 01:52:45 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 11 03:58:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.940	4.001	1955189	1139165	21.385	21.239
2)	SA Decachlor...	10.966	9.266	1897943	1146687	22.966	24.507
Target Compounds							
3)	L1 AR-1016-1	6.264	5.251	1818398	1148864	517.504	525.952
4)	L1 AR-1016-2	6.288	5.271	2524384	1582222	516.476	522.865
5)	L1 AR-1016-3	6.355	5.460	1530021	853203	519.414	523.363
6)	L1 AR-1016-4	6.462	5.512	1284309	685784	521.098	520.042
7)	L1 AR-1016-5	6.778	5.740	1226428	861127	516.547	525.535
8)	L2 AR-1221-1	5.187	4.256	173975	96926	164.199	144.691
9)	L2 AR-1221-2	5.290	4.355	196419	137821	250.056	268.040
10)	L2 AR-1221-3	5.378	4.443	896223	563709	368.414	369.647
11)	L3 AR-1232-1	5.378	4.443	896223	563709	434.633	436.154
12)	L3 AR-1232-2	5.968	5.271	1218915	1582222	1139.988	1190.644
13)	L3 AR-1232-3	6.288	5.460	2524384	853203	1176.350	1172.395
14)	L3 AR-1232-4	6.462	5.557	1284309	805600	1205.941	1319.830
15)	L3 AR-1232-5	6.560	5.740	1010777	861127	1411.726	1288.492
16)	L4 AR-1242-1	6.264	5.251	1818398	1148864	641.583	653.204
17)	L4 AR-1242-2	6.288	5.271	2524384	1582222	648.157	646.932
18)	L4 AR-1242-3	6.355	5.460	1530021	853203	645.018	651.936
19)	L4 AR-1242-4	6.462	5.557	1284309	805600	651.548	643.701
20)	L4 AR-1242-5	7.248	6.117	190203	731179	87.096	428.946 #
21)	L5 AR-1248-1	6.264	5.251	1818398	1148864	836.855	839.096
22)	L5 AR-1248-2	6.560	5.512	1010777	685784	361.697	369.613
23)	L5 AR-1248-3	6.778	5.557	1226428	805600	374.483	429.388
24)	L5 AR-1248-4	7.207	5.740	225893	861127	56.485	382.581 #
25)	L5 AR-1248-5	7.248	6.159	190203	125094	50.295	53.840
26)	L6 AR-1254-1	7.177	6.117	963415	731179	253.812	206.895
27)	L6 AR-1254-2	7.407	6.275	1076421	697038	182.574	223.266
28)	L6 AR-1254-3	7.790	6.711f	581780	1277513	94.951	260.930 #
29)	L6 AR-1254-4	8.085	6.930	422511	192877	80.540	54.997 #
30)	L6 AR-1254-5	8.515	7.356	3545242	2454675	688.326	556.663
31)	L7 AR-1260-1	7.957	6.829	2299420	1688942	544.057	546.354
32)	L7 AR-1260-2	8.224	7.025	3103781	2256624	552.489	556.296
33)	L7 AR-1260-3	8.590	7.179	2124729	1952331	464.815	559.357
34)	L7 AR-1260-4	8.821	7.659	2275139	1420987	480.553	470.565
35)	L7 AR-1260-5	9.148	7.906	4944302	3629555	495.433	485.880
36)	L8 AR-1262-1	8.590	7.356	2124729	2454675	283.820	981.948 #
37)	L8 AR-1262-2	9.148	7.906	4944302	3629555	385.493	412.620
38)	L8 AR-1262-3	9.481f	8.191	3060118	712040	360.700	196.920 #
39)	L8 AR-1262-4	9.535f	8.252	1743900	2363575	521.414	384.129 #
40)	L8 AR-1262-5	10.219	8.748	1189114	732298	263.078	257.220
41)	L9 AR-1268-1	9.481f	8.191	3060118	712040	214.849	73.990 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121120\
 Data File : PO073705.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Dec 2020 1:24
 Operator : DD\AJ
 Sample : L5033-06MSD
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CS-48MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 01:52:45 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 11 03:58:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.535f	8.252	1743900	2363575	141.470	286.890 #
43) L9	AR-1268-3	9.780	8.461	150470	60181	13.864	8.523 #
44) L9	AR-1268-4	10.219	8.748	1189114	732298	253.840	247.397
45) L9	AR-1268-5	10.632	9.025	369184	244460	11.281	12.168



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121120\
Data File : PO073705.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Dec 2020 1:24
Operator : DD\AJ
Sample : L5033-06MSD
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
CS-48MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:52:45 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 11 03:58:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.940 min
Delta R.T.: 0.000 min
Response: 1955189
Conc: 21.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#1 Tetrachloro-m-xylene

R.T.: 4.001 min
Delta R.T.: -0.002 min
Response: 1139165
Conc: 21.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.966 min
Delta R.T.: -0.002 min
Response: 1897943
Conc: 22.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.266 min
Delta R.T.: -0.005 min
Response: 1146687
Conc: 24.51 ng/ml



#3 AR-1016-1


R.T.: 6.264 min
Delta R.T.: 0.000 min
Response: 1818398
Conc: 517.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#3 AR-1016-1

R.T.: 5.251 min
Delta R.T.: -0.002 min
Response: 1148864
Conc: 525.95 ng/ml




#4 AR-1016-2

R.T.: 6.288 min
Delta R.T.: 0.000 min
Response: 2524384
Conc: 516.48 ng/ml



#4 AR-1016-2


R.T.: 5.271 min
Delta R.T.: -0.002 min
Response: 1582222
Conc: 522.87 ng/ml



#5 AR-1016-3


R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 1530021
Conc: 519.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#5 AR-1016-3

R.T.: 5.460 min
Delta R.T.: -0.002 min
Response: 853203
Conc: 523.36 ng/ml




#6 AR-1016-4

R.T.: 6.462 min
Delta R.T.: 0.000 min
Response: 1284309
Conc: 521.10 ng/ml



#6 AR-1016-4


R.T.: 5.512 min
Delta R.T.: -0.002 min
Response: 685784
Conc: 520.04 ng/ml



#7 AR-1016-5


R.T.: 6.778 min
Delta R.T.: 0.000 min
Response: 1226428
Conc: 516.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#7 AR-1016-5

R.T.: 5.740 min
Delta R.T.: -0.002 min
Response: 861127
Conc: 525.54 ng/ml




#8 AR-1221-1

R.T.: 5.187 min
Delta R.T.: -0.004 min
Response: 173975
Conc: 164.20 ng/ml



#8 AR-1221-1


R.T.: 4.256 min
Delta R.T.: -0.003 min
Response: 96926
Conc: 144.69 ng/ml



#9 AR-1221-2


R.T.: 5.290 min
Delta R.T.: 0.000 min
Response: 196419
Conc: 250.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#9 AR-1221-2

R.T.: 4.355 min
Delta R.T.: -0.002 min
Response: 137821
Conc: 268.04 ng/ml




#10 AR-1221-3

R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 896223
Conc: 368.41 ng/ml



#10 AR-1221-3


R.T.: 4.443 min
Delta R.T.: -0.002 min
Response: 563709
Conc: 369.65 ng/ml



#11 AR-1232-1


R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 896223
Conc: 434.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#11 AR-1232-1

R.T.: 4.443 min
Delta R.T.: -0.002 min
Response: 563709
Conc: 436.15 ng/ml




#12 AR-1232-2

R.T.: 5.968 min
Delta R.T.: 0.000 min
Response: 1218915
Conc: 1139.99 ng/ml



#12 AR-1232-2


R.T.: 5.271 min
Delta R.T.: -0.002 min
Response: 1582222
Conc: 1190.64 ng/ml



#13 AR-1232-3


R.T.: 6.288 min
Delta R.T.: 0.000 min
Response: 2524384
Conc: 1176.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#13 AR-1232-3

R.T.: 5.460 min
Delta R.T.: -0.001 min
Response: 853203
Conc: 1172.39 ng/ml




#14 AR-1232-4

R.T.: 6.462 min
Delta R.T.: 0.000 min
Response: 1284309
Conc: 1205.94 ng/ml



#14 AR-1232-4


R.T.: 5.557 min
Delta R.T.: -0.002 min
Response: 805600
Conc: 1319.83 ng/ml



#15 AR-1232-5


R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 1010777
Conc: 1411.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#15 AR-1232-5

R.T.: 5.740 min
Delta R.T.: -0.001 min
Response: 861127
Conc: 1288.49 ng/ml




#16 AR-1242-1

R.T.: 6.264 min
Delta R.T.: -0.002 min
Response: 1818398
Conc: 641.58 ng/ml



#16 AR-1242-1


R.T.: 5.251 min
Delta R.T.: -0.004 min
Response: 1148864
Conc: 653.20 ng/ml



#17 AR-1242-2


R.T.: 6.288 min
Delta R.T.: -0.003 min
Response: 2524384
Conc: 648.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#17 AR-1242-2

R.T.: 5.271 min
Delta R.T.: -0.003 min
Response: 1582222
Conc: 646.93 ng/ml




#18 AR-1242-3

R.T.: 6.355 min
Delta R.T.: -0.002 min
Response: 1530021
Conc: 645.02 ng/ml



#18 AR-1242-3


R.T.: 5.460 min
Delta R.T.: -0.003 min
Response: 853203
Conc: 651.94 ng/ml



#19 AR-1242-4


R.T.: 6.462 min
Delta R.T.: -0.003 min
Response: 1284309
Conc: 651.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#19 AR-1242-4

R.T.: 5.557 min
Delta R.T.: -0.003 min
Response: 805600
Conc: 643.70 ng/ml




#20 AR-1242-5

R.T.: 7.248 min
Delta R.T.: -0.002 min
Response: 190203
Conc: 87.10 ng/ml



#20 AR-1242-5


R.T.: 6.117 min
Delta R.T.: -0.004 min
Response: 731179
Conc: 428.95 ng/ml



#21 AR-1248-1


R.T.: 6.264 min
Delta R.T.: 0.000 min
Response: 1818398
Conc: 836.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#21 AR-1248-1

R.T.: 5.251 min
Delta R.T.: -0.002 min
Response: 1148864
Conc: 839.10 ng/ml




#22 AR-1248-2

R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 1010777
Conc: 361.70 ng/ml



#22 AR-1248-2


R.T.: 5.512 min
Delta R.T.: -0.002 min
Response: 685784
Conc: 369.61 ng/ml



#23 AR-1248-3


R.T.: 6.778 min
Delta R.T.: 0.000 min
Response: 1226428
Conc: 374.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#23 AR-1248-3

R.T.: 5.557 min
Delta R.T.: -0.001 min
Response: 805600
Conc: 429.39 ng/ml




#24 AR-1248-4

R.T.: 7.207 min
Delta R.T.: 0.000 min
Response: 225893
Conc: 56.48 ng/ml



#24 AR-1248-4


R.T.: 5.740 min
Delta R.T.: -0.001 min
Response: 861127
Conc: 382.58 ng/ml



#25 AR-1248-5


R.T.: 7.248 min
Delta R.T.: 0.000 min
Response: 190203
Conc: 50.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#25 AR-1248-5

R.T.: 6.159 min
Delta R.T.: -0.002 min
Response: 125094
Conc: 53.84 ng/ml




#26 AR-1254-1

R.T.: 7.177 min
Delta R.T.: 0.000 min
Response: 963415
Conc: 253.81 ng/ml



#26 AR-1254-1


R.T.: 6.117 min
Delta R.T.: -0.002 min
Response: 731179
Conc: 206.89 ng/ml



#27 AR-1254-2


R.T.: 7.407 min
Delta R.T.: 0.000 min
Response: 1076421
Conc: 182.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#27 AR-1254-2

R.T.: 6.275 min
Delta R.T.: -0.002 min
Response: 697038
Conc: 223.27 ng/ml




#28 AR-1254-3

R.T.: 7.790 min
Delta R.T.: 0.000 min
Response: 581780
Conc: 94.95 ng/ml



#28 AR-1254-3


R.T.: 6.711 min
Delta R.T.: 0.017 min
Response: 1277513
Conc: 260.93 ng/ml



#29 AR-1254-4


R.T.: 8.085 min
Delta R.T.: -0.002 min
Response: 422511
Conc: 80.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#29 AR-1254-4

R.T.: 6.930 min
Delta R.T.: -0.002 min
Response: 192877
Conc: 55.00 ng/ml




#30 AR-1254-5

R.T.: 8.515 min
Delta R.T.: -0.001 min
Response: 3545242
Conc: 688.33 ng/ml



#30 AR-1254-5


R.T.: 7.356 min
Delta R.T.: -0.002 min
Response: 2454675
Conc: 556.66 ng/ml



#31 AR-1260-1


R.T.: 7.957 min
Delta R.T.: -0.001 min
Response: 2299420
Conc: 544.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#31 AR-1260-1

R.T.: 6.829 min
Delta R.T.: -0.002 min
Response: 1688942
Conc: 546.35 ng/ml




#32 AR-1260-2

R.T.: 8.224 min
Delta R.T.: -0.001 min
Response: 3103781
Conc: 552.49 ng/ml



#32 AR-1260-2


R.T.: 7.025 min
Delta R.T.: -0.003 min
Response: 2256624
Conc: 556.30 ng/ml



#33 AR-1260-3


R.T.: 8.590 min
Delta R.T.: 0.000 min
Response: 2124729
Conc: 464.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#33 AR-1260-3

R.T.: 7.179 min
Delta R.T.: -0.003 min
Response: 1952331
Conc: 559.36 ng/ml




#34 AR-1260-4

R.T.: 8.821 min
Delta R.T.: -0.002 min
Response: 2275139
Conc: 480.55 ng/ml



#34 AR-1260-4


R.T.: 7.659 min
Delta R.T.: -0.003 min
Response: 1420987
Conc: 470.57 ng/ml



#35 AR-1260-5


R.T.: 9.148 min
Delta R.T.: -0.002 min
Response: 4944302
Conc: 495.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#35 AR-1260-5

R.T.: 7.906 min
Delta R.T.: -0.003 min
Response: 3629555
Conc: 485.88 ng/ml




#36 AR-1262-1

R.T.: 8.590 min
Delta R.T.: 0.000 min
Response: 2124729
Conc: 283.82 ng/ml



#36 AR-1262-1


R.T.: 7.356 min
Delta R.T.: -0.001 min
Response: 2454675
Conc: 981.95 ng/ml



#37 AR-1262-2


R.T.: 9.148 min
Delta R.T.: 0.000 min
Response: 4944302
Conc: 385.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#37 AR-1262-2

R.T.: 7.906 min
Delta R.T.: -0.001 min
Response: 3629555
Conc: 412.62 ng/ml




#38 AR-1262-3

R.T.: 9.481 min
Delta R.T.: 0.022 min
Response: 3060118
Conc: 360.70 ng/ml



#38 AR-1262-3


R.T.: 8.191 min
Delta R.T.: -0.002 min
Response: 712040
Conc: 196.92 ng/ml



#39 AR-1262-4


R.T.: 9.535 min
Delta R.T.: -0.018 min
Response: 1743900
Conc: 521.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#39 AR-1262-4

R.T.: 8.252 min
Delta R.T.: -0.004 min
Response: 2363575
Conc: 384.13 ng/ml




#40 AR-1262-5

R.T.: 10.219 min
Delta R.T.: -0.001 min
Response: 1189114
Conc: 263.08 ng/ml



#40 AR-1262-5


R.T.: 8.748 min
Delta R.T.: -0.002 min
Response: 732298
Conc: 257.22 ng/ml



#41 AR-1268-1


R.T.: 9.481 min
Delta R.T.: 0.022 min
Response: 3060118
Conc: 214.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#41 AR-1268-1

R.T.: 8.191 min
Delta R.T.: -0.001 min
Response: 712040
Conc: 73.99 ng/ml




#42 AR-1268-2

R.T.: 9.535 min
Delta R.T.: -0.021 min
Response: 1743900
Conc: 141.47 ng/ml



#42 AR-1268-2


R.T.: 8.252 min
Delta R.T.: -0.005 min
Response: 2363575
Conc: 286.89 ng/ml



#43 AR-1268-3


R.T.: 9.780 min
Delta R.T.: -0.001 min
Response: 150470
Conc: 13.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD




#43 AR-1268-3

R.T.: 8.461 min
Delta R.T.: -0.002 min
Response: 60181
Conc: 8.52 ng/ml




#44 AR-1268-4

R.T.: 10.219 min
Delta R.T.: 0.000 min
Response: 1189114
Conc: 253.84 ng/ml



#44 AR-1268-4


R.T.: 8.748 min
Delta R.T.: 0.000 min
Response: 732298
Conc: 247.40 ng/ml



#45 AR-1268-5

R.T.: 10.632 min
Delta R.T.: 0.000 min
Response: 369184
Conc: 11.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-48MSD



#45 AR-1268-5

R.T.: 9.025 min
Delta R.T.: -0.003 min
Response: 244460
Conc: 12.17 ng/ml