

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121123\
 Data File : PO100432.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2023 04:04
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 04:57:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 11 00:11:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.472	3.699	113.4E6	101.1E6	45.300	46.441
2)	SA Decachlor...	10.341	8.782	76488290	76162551	54.803	47.204
Target Compounds							
26)	L6 AR-1254-1	6.524	5.614	36741828	32958585	431.955	419.429
27)	L6 AR-1254-2	6.746	5.763	54733931	28996152	440.561	411.802
28)	L6 AR-1254-3	7.116	6.171	55277611	49623617	462.848	409.840
29)	L6 AR-1254-4	7.404	6.401	34775335	28828500	501.462	439.010
30)	L6 AR-1254-5	7.828	6.823	39546741	45834703	482.472	415.351

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121123\
Data File : PO100432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2023 04:04
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 04:57:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 11 00:11:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

