

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121123\
 Data File : PO100414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2023 19:48
 Operator : YP/AJ
 Sample : 05741-03 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC-271246-

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/12/2023
 Supervised By :Ankita Jodhani 12/12/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 23:57:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 11 00:11:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.472	3.699	2058158	2511805	0.822m	1.154m#
2) SA Decachlor...	10.344	8.785	2133345	1887107	1.529	1.170m
Target Compounds						
26) L6 AR-1254-1	6.527	5.615	1367776	1180359	16.080	15.021m
27) L6 AR-1254-2	6.747	5.763	1825916	1154951	14.697	16.403m
28) L6 AR-1254-3	7.116	6.172	1820819	2123248	15.246m	17.536m
29) L6 AR-1254-4	7.406	6.403	1159516	1110855	16.720	16.916
30) L6 AR-1254-5	7.830	6.823	1444630	2138936	17.625m	19.383

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121123\
Data File : PO100414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2023 19:48
Operator : YP/AJ
Sample : 05741-03 10X
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC-271246-


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/12/2023
Supervised By :Ankita Jodhani 12/12/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 23:57:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 11 00:11:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.472 min
Delta R.T.: 0.000 min
Response: 2058158
Conc: 0.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC-271246-


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/12/2023
Supervised By :Ankita Jodhani 12/12/2023




#1 Tetrachloro-m-xylene

R.T.: 3.699 min
Delta R.T.: 0.000 min
Response: 2511805
Conc: 1.15 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.344 min
Delta R.T.: -0.002 min
Response: 2133345
Conc: 1.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.785 min
Delta R.T.: -0.002 min
Response: 1887107
Conc: 1.17 ng/ml m




#26 AR-1254-1

R.T.: 6.527 min
Delta R.T.: 0.001 min
Response: 1367776
Conc: 16.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC-271246-


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/12/2023
Supervised By :Ankita Jodhani 12/12/2023




#26 AR-1254-1

R.T.: 5.615 min
Delta R.T.: -0.001 min
Response: 1180359
Conc: 15.02 ng/ml m




#27 AR-1254-2

R.T.: 6.747 min
Delta R.T.: 0.000 min
Response: 1825916
Conc: 14.70 ng/ml



#27 AR-1254-2

R.T.: 5.763 min
Delta R.T.: 0.000 min
Response: 1154951
Conc: 16.40 ng/ml m




#28 AR-1254-3

R.T.: 7.116 min
Delta R.T.: 0.000 min
Response: 1820819
Conc: 15.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC-271246-


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/12/2023
Supervised By :Ankita Jodhani 12/12/2023




#28 AR-1254-3

R.T.: 6.172 min
Delta R.T.: -0.001 min
Response: 2123248
Conc: 17.54 ng/ml m




#29 AR-1254-4

R.T.: 7.406 min
Delta R.T.: 0.000 min
Response: 1159516
Conc: 16.72 ng/ml



#29 AR-1254-4

R.T.: 6.403 min
Delta R.T.: 0.000 min
Response: 1110855
Conc: 16.92 ng/ml




#30 AR-1254-5

R.T.: 7.830 min
Delta R.T.: 0.000 min
Response: 1444630
Conc: 17.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC-271246-

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/12/2023
Supervised By :Ankita Jodhani 12/12/2023



#30 AR-1254-5

R.T.: 6.823 min
Delta R.T.: -0.002 min
Response: 2138936
Conc: 19.38 ng/ml