

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121124\
 Data File : P0108517.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2024 22:01
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 01:20:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 05:58:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.708	3.705	471.4E6	284.4E6	54.179	56.108
2) SA Decachlor...	8.783	8.731	344.7E6	214.9E6	47.191	55.455
Target Compounds						
21) L5 AR-1248-1	4.806	4.794	102.8E6	58160892	513.335	553.287
22) L5 AR-1248-2	5.044	5.030	132.3E6	73960291	477.189	495.289
23) L5 AR-1248-3	5.261	5.073	176.9E6	85722806	518.863	541.140
24) L5 AR-1248-4	5.616	5.244	243.1E6	88423066	510.247	484.376
25) L5 AR-1248-5	5.658	5.638	170.1E6	94543274	502.224	538.391



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121124\
Data File : PO108517.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2024 22:01
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:20:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 05:58:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.708 min
Delta R.T.: -0.002 min
Response: 471363004
Conc: 54.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.705 min
Delta R.T.: -0.003 min
Response: 284433984
Conc: 56.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.783 min
Delta R.T.: -0.008 min
Response: 344688637
Conc: 47.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.731 min
Delta R.T.: -0.011 min
Response: 214883742
Conc: 55.46 ng/ml



#21 AR-1248-1


R.T.: 4.806 min
Delta R.T.: -0.002 min
Response: 102807232
Conc: 513.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.794 min
Delta R.T.: -0.004 min
Response: 58160892
Conc: 553.29 ng/ml




#22 AR-1248-2

R.T.: 5.044 min
Delta R.T.: -0.003 min
Response: 132269901
Conc: 477.19 ng/ml



#22 AR-1248-2


R.T.: 5.030 min
Delta R.T.: -0.005 min
Response: 73960291
Conc: 495.29 ng/ml



#23 AR-1248-3


R.T.: 5.261 min
Delta R.T.: -0.002 min
Response: 176932241
Conc: 518.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.073 min
Delta R.T.: -0.004 min
Response: 85722806
Conc: 541.14 ng/ml




#24 AR-1248-4

R.T.: 5.616 min
Delta R.T.: -0.002 min
Response: 243146974
Conc: 510.25 ng/ml



#24 AR-1248-4


R.T.: 5.244 min
Delta R.T.: -0.004 min
Response: 88423066
Conc: 484.38 ng/ml



#25 AR-1248-5

R.T.: 5.658 min
Delta R.T.: -0.002 min
Response: 170058563
Conc: 502.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.638 min
Delta R.T.: -0.005 min
Response: 94543274
Conc: 538.39 ng/ml