

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121222\
 Data File : P0091265.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2022 21:40
 Operator : YP/AJ
 Sample : N5999-01MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MH-20MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 23:31:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.296	3.474	67399199	21737086	19.934	20.711
2) SA Decachlor...	10.041	8.474	45782796	19653944	17.236	20.244
Target Compounds						
3) L1 AR-1016-1	5.472	4.558	49794769	17839627	442.491	484.089
4) L1 AR-1016-2	5.495	4.577	71460557	24802673	439.545	481.332
5) L1 AR-1016-3	5.557	4.752	44531791	13560904	427.124	498.280
6) L1 AR-1016-4	5.656	4.796	34961670	11568877	431.453	500.482
7) L1 AR-1016-5	5.953	5.008	36273849	14361352	413.638	486.881
31) L7 AR-1260-1	7.087	6.046	62053654	27048711	404.504	467.145
32) L7 AR-1260-2	7.346	6.236	71123948	31788386	407.487	465.518
33) L7 AR-1260-3	7.708	6.388	46920127	29629820	412.761	451.722
34) L7 AR-1260-4	7.934	6.862	57255821	21956623	422.525	459.728
35) L7 AR-1260-5	8.251	7.107	99533864	48561975	391.863	446.702


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121222\
Data File : P0091265.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2022 21:40
Operator : YP/AJ
Sample : N5999-01MS
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MH-20MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 23:31:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.296 min
Delta R.T.: -0.002 min
Response: 67399199
Conc: 19.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-20MS




#1 Tetrachloro-m-xylene

R.T.: 3.474 min
Delta R.T.: -0.003 min
Response: 21737086
Conc: 20.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.041 min
Delta R.T.: -0.009 min
Response: 45782796
Conc: 17.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.474 min
Delta R.T.: -0.008 min
Response: 19653944
Conc: 20.24 ng/ml



#3 AR-1016-1


R.T.: 5.472 min
Delta R.T.: -0.003 min
Response: 49794769
Conc: 442.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-20MS




#3 AR-1016-1

R.T.: 4.558 min
Delta R.T.: -0.004 min
Response: 17839627
Conc: 484.09 ng/ml




#4 AR-1016-2

R.T.: 5.495 min
Delta R.T.: -0.003 min
Response: 71460557
Conc: 439.54 ng/ml



#4 AR-1016-2


R.T.: 4.577 min
Delta R.T.: -0.004 min
Response: 24802673
Conc: 481.33 ng/ml



#5 AR-1016-3


R.T.: 5.557 min
Delta R.T.: -0.003 min
Response: 44531791
Conc: 427.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-20MS




#5 AR-1016-3

R.T.: 4.752 min
Delta R.T.: -0.005 min
Response: 13560904
Conc: 498.28 ng/ml




#6 AR-1016-4

R.T.: 5.656 min
Delta R.T.: -0.003 min
Response: 34961670
Conc: 431.45 ng/ml



#6 AR-1016-4


R.T.: 4.796 min
Delta R.T.: -0.005 min
Response: 11568877
Conc: 500.48 ng/ml



#7 AR-1016-5


R.T.: 5.953 min
Delta R.T.: -0.004 min
Response: 36273849
Conc: 413.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-20MS




#7 AR-1016-5

R.T.: 5.008 min
Delta R.T.: -0.005 min
Response: 14361352
Conc: 486.88 ng/ml




#31 AR-1260-1

R.T.: 7.087 min
Delta R.T.: -0.005 min
Response: 62053654
Conc: 404.50 ng/ml



#31 AR-1260-1


R.T.: 6.046 min
Delta R.T.: -0.006 min
Response: 27048711
Conc: 467.14 ng/ml



#32 AR-1260-2


R.T.: 7.346 min
Delta R.T.: -0.005 min
Response: 71123948
Conc: 407.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-20MS




#32 AR-1260-2

R.T.: 6.236 min
Delta R.T.: -0.006 min
Response: 31788386
Conc: 465.52 ng/ml




#33 AR-1260-3

R.T.: 7.708 min
Delta R.T.: -0.005 min
Response: 46920127
Conc: 412.76 ng/ml



#33 AR-1260-3


R.T.: 6.388 min
Delta R.T.: -0.006 min
Response: 29629820
Conc: 451.72 ng/ml



#34 AR-1260-4


R.T.: 7.934 min
Delta R.T.: -0.005 min
Response: 57255821
Conc: 422.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-20MS




#34 AR-1260-4

R.T.: 6.862 min
Delta R.T.: -0.007 min
Response: 21956623
Conc: 459.73 ng/ml



#35 AR-1260-5

R.T.: 8.251 min
Delta R.T.: -0.005 min
Response: 99533864
Conc: 391.86 ng/ml



#35 AR-1260-5

R.T.: 7.107 min
Delta R.T.: -0.007 min
Response: 48561975
Conc: 446.70 ng/ml