

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO121318\
 Data File : PO052485.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Dec 2018 15:58
 Operator : SM/SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 01:31:24 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO120318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 04 03:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.261	3.282	23131677	5872019	26.869	26.328
2) SA Decachlor...	0.000	7.970	0	4630838	N.D.	22.213 #



Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO121318\
Data File : PO052485.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Dec 2018 15:58
Operator : SM/SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 01:31:24 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO120318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 04 03:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped


Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.261 min
Delta R.T.: -0.022 min
Response: 23131677
Conc: 26.87 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 3.282 min
Delta R.T.: -0.014 min
Response: 5872019
Conc: 26.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.872 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 7.970 min
Delta R.T.: -0.030 min
Response: 4630838
Conc: 22.21 ng/ml