

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121318\
 Data File : P0052478.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Dec 2018 14:00
 Operator : SM/SJ
 Sample : PB115628BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB115628BL

Manual Integrations
 APPROVED

Sohil
 12/14/2018 4:08:14 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 01:29:34 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 04 03:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.275	3.283	18981025	4350489	22.047	19.506
2) SA Decachlor...	9.840	7.975	8700044	3849269	17.384m	18.464

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121318\
Data File : PO052478.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Dec 2018 14:00
Operator : SM/SJ
Sample : PB115628BL
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB115628BL


Manual Integrations
APPROVED

Sohil
12/14/2018 4:08:14 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 01:29:34 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 04 03:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.275 min
Delta R.T.: -0.008 min
Response: 18981025
Conc: 22.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB115628BL


Manual Integrations
APPROVED

Sohil
12/14/2018 4:08:14 PM




#1 Tetrachloro-m-xylene

R.T.: 3.283 min
Delta R.T.: -0.013 min
Response: 4350489
Conc: 19.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.840 min
Delta R.T.: -0.032 min
Response: 8700044
Conc: 17.38 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.975 min
Delta R.T.: -0.025 min
Response: 3849269
Conc: 18.46 ng/ml