

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121318\
 Data File : P0052488.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Dec 2018 16:49
 Operator : SM/SJ
 Sample : J6319-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 061_SW-N-1

Manual Integrations
 APPROVED

Sohil
 12/14/2018 4:08:33 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 01:32:01 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 04 03:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.262	3.282	21274146	5234229	24.711	23.469
2)	SA Decachlor...	9.821	7.971	7994443	3445814	15.975m	16.529

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121318\
Data File : PO052488.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Dec 2018 16:49
Operator : SM/SJ
Sample : J6319-07
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
061_SW-N-1


Manual Integrations
APPROVED

Sohil
12/14/2018 4:08:33 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 01:32:01 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 04 03:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.262 min
Delta R.T.: -0.021 min
Response: 21274146
Conc: 24.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
061_SW-N-1


Manual Integrations
APPROVED

Sohil
12/14/2018 4:08:33 PM




#1 Tetrachloro-m-xylene

R.T.: 3.282 min
Delta R.T.: -0.014 min
Response: 5234229
Conc: 23.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.821 min
Delta R.T.: -0.050 min
Response: 7994443
Conc: 15.97 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.971 min
Delta R.T.: -0.029 min
Response: 3445814
Conc: 16.53 ng/ml