

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121318\
 Data File : P0052506.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Dec 2018 21:54
 Operator : SM/SJ
 Sample : J6353-06
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P001-SS003-0002-01

Manual Integrations
 APPROVED

Sohil
 12/14/2018 4:09:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 01:39:28 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 04 03:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.260	3.280	16025396	4322847	18.614	19.382
2) SA Decachlor...	9.817	7.970	6774981	3348096	13.538m	16.060

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121318\
Data File : PO052506.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Dec 2018 21:54
Operator : SM/SJ
Sample : J6353-06
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
P001-SS003-0002-01


Manual Integrations
APPROVED

Sohil
12/14/2018 4:09:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 01:39:28 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 04 03:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.260 min
Delta R.T.: -0.023 min
Response: 16025396
Conc: 18.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS003-0002-01


Manual Integrations
APPROVED

Sohil
12/14/2018 4:09:04 PM




#1 Tetrachloro-m-xylene

R.T.: 3.280 min
Delta R.T.: -0.016 min
Response: 4322847
Conc: 19.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.817 min
Delta R.T.: -0.055 min
Response: 6774981
Conc: 13.54 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.970 min
Delta R.T.: -0.030 min
Response: 3348096
Conc: 16.06 ng/ml