Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121322\
Data File : PO091332.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2022 00:43
Operator : YP/AJ
Sample : N6012-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PCB12622-01

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 12/14/2022
Supervised By :Ankita Jodhani 12/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 03:45:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.295	3.472	69880360	24369788	20.668	23.219	
2)	SA Decachlor...	10.042	8.472	52511166	22729794	19.769	23.412	
	Target Compounds							
26)	L6 AR-1254-1	6.333	5.360	3403557	1556482	27.100	26.558	
27)	L6 AR-1254-2	6.552	5.509	6425086	1864202	34.075	34.930	
28)	L6 AR-1254-3	6.921	5.913	8284950	4361726	44.201	53.537	
29)	L6 AR-1254-4	7.208	6.143	4723395	1819855	34.193	37.195	
30)	L6 AR-1254-5	7.631	6.562	8253215	5717140	53.792	76.694	#
31)	L7 AR-1260-1	7.086	6.044	4992237	2575724	32.542	44.484	#
32)	L7 AR-1260-2	7.344	6.235	7786629	2624684	44.612	38.437	
33)	L7 AR-1260-3	7.707	6.386	4252171	2681226	37.407	40.877	
34)	L7 AR-1260-4	7.931	6.860	5975540	1680885	44.097m	35.194	
35)	L7 AR-1260-5	8.250	7.105	8698603	4311334	34.246	39.658	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121322\
Data File : PO091332.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2022 00:43
Operator : YP/AJ
Sample : N6012-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PCB12622-01

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 12/14/2022
Supervised By :Ankita Jodhani 12/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 03:45:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm