

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121322\
 Data File : PO091315.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2022 19:25
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 20:57:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.296	3.473	153.5E6	53170112	45.386	50.659
2) SA Decachlor...	10.042	8.473	110.9E6	47288039	41.753	48.707
Target Compounds						
26) L6 AR-1254-1	6.333	5.361	52554508	28058273	418.457	478.754
27) L6 AR-1254-2	6.553	5.510	79078010	25842326	419.384	484.208
28) L6 AR-1254-3	6.921	5.914	80584386	38442270	429.927	471.848
29) L6 AR-1254-4	7.209	6.144	59629940	23340500	431.664	477.045
30) L6 AR-1254-5	7.631	6.564	63922945	35054515	416.630	470.249



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121322\
Data File : PO091315.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2022 19:25
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 20:57:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.296 min
Delta R.T.: -0.002 min
Response: 153454523
Conc: 45.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.473 min
Delta R.T.: -0.003 min
Response: 53170112
Conc: 50.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.042 min
Delta R.T.: -0.008 min
Response: 110904183
Conc: 41.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.473 min
Delta R.T.: -0.009 min
Response: 47288039
Conc: 48.71 ng/ml



#26 AR-1254-1


R.T.: 6.333 min
Delta R.T.: -0.005 min
Response: 52554508
Conc: 418.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.361 min
Delta R.T.: -0.007 min
Response: 28058273
Conc: 478.75 ng/ml




#27 AR-1254-2

R.T.: 6.553 min
Delta R.T.: -0.006 min
Response: 79078010
Conc: 419.38 ng/ml



#27 AR-1254-2


R.T.: 5.510 min
Delta R.T.: -0.007 min
Response: 25842326
Conc: 484.21 ng/ml



#28 AR-1254-3


R.T.: 6.921 min
Delta R.T.: -0.006 min
Response: 80584386
Conc: 429.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.914 min
Delta R.T.: -0.007 min
Response: 38442270
Conc: 471.85 ng/ml




#29 AR-1254-4

R.T.: 7.209 min
Delta R.T.: -0.007 min
Response: 59629940
Conc: 431.66 ng/ml



#29 AR-1254-4


R.T.: 6.144 min
Delta R.T.: -0.008 min
Response: 23340500
Conc: 477.04 ng/ml



#30 AR-1254-5

R.T.: 7.631 min
Delta R.T.: -0.007 min
Response: 63922945
Conc: 416.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.564 min
Delta R.T.: -0.008 min
Response: 35054515
Conc: 470.25 ng/ml