

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121323\
 Data File : PO100451.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2023 23:59
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 01:51:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 14 01:42:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.468	3.694	38856730	37970812	24.609	25.769
2) SA Decachlor...	10.327	8.773	24055725	27144513	24.986	25.134
Target Compounds						
16) L4 AR-1242-1	5.653	4.797	8226992	6263992	252.474	269.787
17) L4 AR-1242-2	5.675	4.816	12144910	8743496	246.944	267.531
18) L4 AR-1242-3	5.738	4.995	7483106	4592291	244.645	265.642
19) L4 AR-1242-4	5.838	5.079	5706853	5035218	240.795	269.573
20) L4 AR-1242-5	6.584	5.609	5238217	6151271	242.249	282.571


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121323\
Data File : PO100451.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2023 23:59
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 01:51:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121323.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 14 01:42:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.468 min
Delta R.T.: 0.000 min
Response: 38856730
Conc: 24.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 37970812
Conc: 25.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.327 min
Delta R.T.: 0.000 min
Response: 24055725
Conc: 24.99 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.773 min
Delta R.T.: 0.000 min
Response: 27144513
Conc: 25.13 ng/ml



#16 AR-1242-1


R.T.: 5.653 min
Delta R.T.: 0.000 min
Response: 8226992
Conc: 252.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.797 min
Delta R.T.: 0.000 min
Response: 6263992
Conc: 269.79 ng/ml




#17 AR-1242-2

R.T.: 5.675 min
Delta R.T.: 0.000 min
Response: 12144910
Conc: 246.94 ng/ml



#17 AR-1242-2


R.T.: 4.816 min
Delta R.T.: 0.000 min
Response: 8743496
Conc: 267.53 ng/ml



#18 AR-1242-3


R.T.: 5.738 min
Delta R.T.: 0.000 min
Response: 7483106
Conc: 244.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.995 min
Delta R.T.: 0.000 min
Response: 4592291
Conc: 265.64 ng/ml




#19 AR-1242-4

R.T.: 5.838 min
Delta R.T.: 0.000 min
Response: 5706853
Conc: 240.79 ng/ml



#19 AR-1242-4


R.T.: 5.079 min
Delta R.T.: 0.000 min
Response: 5035218
Conc: 269.57 ng/ml



#20 AR-1242-5

R.T.: 6.584 min
Delta R.T.: 0.000 min
Response: 5238217
Conc: 242.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 6151271
Conc: 282.57 ng/ml