

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121323\
 Data File : PO100470.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2023 05:19
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO121323AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 05:33:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 14 05:06:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.468	3.695	77510438	74337451	50.646	50.287
2)	SA Decachlor...	10.325	8.772	41459912	50353211	43.988	47.171
Target Compounds							
16)	L4 AR-1242-1	5.653	4.797	15140285	11439127	478.033	488.732
17)	L4 AR-1242-2	5.675	4.816	22864245	15591749	477.432	478.250
18)	L4 AR-1242-3	5.738	4.995	14544416	8508935	488.289	488.869
19)	L4 AR-1242-4	5.837	5.078	11059091	9128743	493.513	484.598
20)	L4 AR-1242-5	6.582	5.608	9740861	10008737	465.229	455.379

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121323\
Data File : PO100470.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2023 05:19
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO121323AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 05:33:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121323.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 14 05:06:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

