

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121420\
 Data File : PO073773.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Dec 2020 18:20
 Operator : DD\AJ
 Sample : L5026-26
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-61(4.5-5.0)

Manual Integrations
 APPROVED

Ankita
 12/15/2020 1:09:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 03:35:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 11 03:58:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.939	3.998	2042302	1252313	22.338	23.348
2) SA Decachlor...	10.965	9.264	1837918	1103981	22.240	23.594
Target Compounds						
26) L6 AR-1254-1	7.175	6.113	69107	59938	18.206	16.960m
27) L6 AR-1254-2	7.407	6.271	107347	55141	18.207	17.662
28) L6 AR-1254-3	7.789	6.688	118183	97028	19.288	19.818
29) L6 AR-1254-4	8.087	6.927	137592	84515	26.228	24.099
30) L6 AR-1254-5	8.515	7.354	121160	83877	23.524	19.021

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121420\
Data File : P0073773.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Dec 2020 18:20
Operator : DD\AJ
Sample : L5026-26
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SB-61(4.5-5.0)


Manual Integrations
APPROVED

Ankita
12/15/2020 1:09:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 03:35:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 11 03:58:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.939 min
Delta R.T.: -0.002 min
Response: 2042302
Conc: 22.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-61(4.5-5.0)


Manual Integrations
APPROVED

Ankita
12/15/2020 1:09:39 PM




#1 Tetrachloro-m-xylene

R.T.: 3.998 min
Delta R.T.: -0.005 min
Response: 1252313
Conc: 23.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.965 min
Delta R.T.: -0.003 min
Response: 1837918
Conc: 22.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.264 min
Delta R.T.: -0.007 min
Response: 1103981
Conc: 23.59 ng/ml




#26 AR-1254-1

R.T.: 7.175 min
Delta R.T.: -0.002 min
Response: 69107
Conc: 18.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-61(4.5-5.0)


Manual Integrations
APPROVED

Ankita
12/15/2020 1:09:39 PM




#26 AR-1254-1

R.T.: 6.113 min
Delta R.T.: -0.005 min
Response: 59938
Conc: 16.96 ng/ml m




#27 AR-1254-2

R.T.: 7.407 min
Delta R.T.: 0.000 min
Response: 107347
Conc: 18.21 ng/ml



#27 AR-1254-2

R.T.: 6.271 min
Delta R.T.: -0.006 min
Response: 55141
Conc: 17.66 ng/ml




#28 AR-1254-3

R.T.: 7.789 min
Delta R.T.: -0.001 min
Response: 118183
Conc: 19.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-61(4.5-5.0)


Manual Integrations
APPROVED

Ankita
12/15/2020 1:09:39 PM




#28 AR-1254-3

R.T.: 6.688 min
Delta R.T.: -0.006 min
Response: 97028
Conc: 19.82 ng/ml




#29 AR-1254-4

R.T.: 8.087 min
Delta R.T.: 0.000 min
Response: 137592
Conc: 26.23 ng/ml



#29 AR-1254-4

R.T.: 6.927 min
Delta R.T.: -0.005 min
Response: 84515
Conc: 24.10 ng/ml




#30 AR-1254-5

R.T.: 8.515 min
Delta R.T.: -0.002 min
Response: 121160
Conc: 23.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-61(4.5-5.0)

Manual Integrations
APPROVED

Ankita
12/15/2020 1:09:39 PM



#30 AR-1254-5

R.T.: 7.354 min
Delta R.T.: -0.004 min
Response: 83877
Conc: 19.02 ng/ml