

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO121421\  
 Data File : PO083692.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 15 Dec 2021 00:23  
 Operator : AJ\MA  
 Sample : M5077-06 10X  
 Misc :  
 ALS Vial : 34 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 06-A-06-B-06-C

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Dec 15 02:53:39 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO120621.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Tue Dec 07 13:18:47 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1    | RT#2  | Resp#1 | Resp#2 | ng/ml | ng/ml |
|-----------------------------|-----------------|---------|-------|--------|--------|-------|-------|
| -----                       |                 |         |       |        |        |       |       |
| System Monitoring Compounds |                 |         |       |        |        |       |       |
| 1)                          | SA Tetrachlo... | 4.934   | 3.940 | 162708 | 62781  | 1.705 | 2.018 |
| 2)                          | SA Decachlor... | 10.999f | 9.233 | 142575 | 64598  | 1.996 | 1.980 |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO121421\  
Data File : PO083692.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 15 Dec 2021 00:23  
Operator : AJ\MA  
Sample : M5077-06 10X  
Misc :  
ALS Vial : 34 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
06-A-06-B-06-C

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Dec 15 02:53:39 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO120621.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Tue Dec 07 13:18:47 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

