

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121422\
 Data File : PO091354.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2022 16:09
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 08:12:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.297	3.473	162.6E6	56165808	48.100	53.514
2) SA Decachlor...	10.044	8.473	118.9E6	50326960	44.781	51.837
Target Compounds						
21) L5 AR-1248-1	5.473	4.558	32724953	11920321	459.406	506.692
22) L5 AR-1248-2	5.749	4.795	44802360	16461802	422.627	502.805
23) L5 AR-1248-3	5.953	4.837	50793337	17104017	431.617	503.951
24) L5 AR-1248-4	6.360	5.007	53496956	19940932	437.467	499.744
25) L5 AR-1248-5	6.398	5.402	51863425	18264756	452.179	509.548



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121422\
Data File : PO091354.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2022 16:09
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 08:12:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.297 min
Delta R.T.: -0.001 min
Response: 162630090
Conc: 48.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.473 min
Delta R.T.: -0.003 min
Response: 56165808
Conc: 53.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.044 min
Delta R.T.: -0.006 min
Response: 118947907
Conc: 44.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.473 min
Delta R.T.: -0.008 min
Response: 50326960
Conc: 51.84 ng/ml



#21 AR-1248-1


R.T.: 5.473 min
Delta R.T.: -0.003 min
Response: 32724953
Conc: 459.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.558 min
Delta R.T.: -0.004 min
Response: 11920321
Conc: 506.69 ng/ml




#22 AR-1248-2

R.T.: 5.749 min
Delta R.T.: -0.003 min
Response: 44802360
Conc: 422.63 ng/ml



#22 AR-1248-2


R.T.: 4.795 min
Delta R.T.: -0.005 min
Response: 16461802
Conc: 502.81 ng/ml



#23 AR-1248-3


R.T.: 5.953 min
Delta R.T.: -0.004 min
Response: 50793337
Conc: 431.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.837 min
Delta R.T.: -0.004 min
Response: 17104017
Conc: 503.95 ng/ml




#24 AR-1248-4

R.T.: 6.360 min
Delta R.T.: -0.003 min
Response: 53496956
Conc: 437.47 ng/ml



#24 AR-1248-4


R.T.: 5.007 min
Delta R.T.: -0.005 min
Response: 19940932
Conc: 499.74 ng/ml



#25 AR-1248-5

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 51863425
Conc: 452.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.402 min
Delta R.T.: -0.005 min
Response: 18264756
Conc: 509.55 ng/ml