

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121521\
 Data File : PO083710.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Dec 2021 18:30
 Operator : AJ\MA
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 03:52:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 16 03:52:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.932	3.938	4932399	1472063	50.000	50.000
2)	SA Decachlor...	10.999	9.231	3467632	1422818	50.000	50.000
Target Compounds							
11)	L3 AR-1232-1	5.373	4.381	973472	371894	500.000	500.000
12)	L3 AR-1232-2	5.967	5.211	509136	358238	500.000	500.000
13)	L3 AR-1232-3	6.288	5.402	913166	185101	500.000	500.000
14)	L3 AR-1232-4	6.462	5.499	448893	162448	500.000	500.000
15)	L3 AR-1232-5	6.560	5.683	301973	174723	500.000	500.000
36)	L8 AR-1262-1	8.601	0.000	19788	0	4.906	N.D. #
39)	L8 AR-1262-4	9.594f	8.214	80749	15336	40.598	5.177 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121521\
Data File : PO083710.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Dec 2021 18:30
Operator : AJ\MA
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 03:52:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 16 03:52:31 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.932 min
Delta R.T.: 0.000 min
Response: 4932399
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.938 min
Delta R.T.: 0.000 min
Response: 1472063
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.999 min
Delta R.T.: 0.000 min
Response: 3467632
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.231 min
Delta R.T.: 0.000 min
Response: 1422818
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 5.373 min
Delta R.T.: 0.000 min
Response: 973472
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.381 min
Delta R.T.: 0.000 min
Response: 371894
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 5.967 min
Delta R.T.: 0.000 min
Response: 509136
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 5.211 min
Delta R.T.: 0.000 min
Response: 358238
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 6.288 min
Delta R.T.: 0.000 min
Response: 913166
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 5.402 min
Delta R.T.: 0.000 min
Response: 185101
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 6.462 min
Delta R.T.: 0.000 min
Response: 448893
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.499 min
Delta R.T.: 0.000 min
Response: 162448
Conc: 500.00 ng/ml



#15 AR-1232-5


R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 301973
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#15 AR-1232-5

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 174723
Conc: 500.00 ng/ml




#36 AR-1262-1

R.T.: 8.601 min
Delta R.T.: 0.005 min
Response: 19788
Conc: 4.91 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 7.312 min
Response: 0
Conc: N.D.



#39 AR-1262-4

R.T.: 9.594 min
Delta R.T.: 0.026 min
Response: 80749
Conc: 40.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500



#39 AR-1262-4

R.T.: 8.214 min
Delta R.T.: -0.003 min
Response: 15336
Conc: 5.18 ng/ml