

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121521\
 Data File : PO083708.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Dec 2021 17:56
 Operator : AJ\MA
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 01:38:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 16 01:19:03 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.933	3.938	432384	144015	4.539	4.910
2) SA Decachlor...	11.000	9.232	338420	140242	4.941	4.994
Target Compounds						
3) L1 AR-1016-1	6.265	5.191	161655	66519	50.795	54.062
4) L1 AR-1016-2	6.288	5.213	233206	94241	51.243	54.881
5) L1 AR-1016-3	6.355	5.402	134545	47360	49.464	52.033
6) L1 AR-1016-4	6.462	5.456	111893	35393	50.388	47.914
7) L1 AR-1016-5	6.778	5.683	103231	46104	48.614	50.848
31) L7 AR-1260-1	7.961	6.780	157332	99928	48.556	59.620
32) L7 AR-1260-2	8.229	6.979	193293	109735	49.903	53.145
33) L7 AR-1260-3	8.594	7.132	152490	92155	51.476	52.057
34) L7 AR-1260-4	8.826	7.616	174433	69469	50.994	47.147
35) L7 AR-1260-5	9.157	7.865	340629	184071	50.556	47.563


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121521\
Data File : PO083708.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Dec 2021 17:56
Operator : AJ\MA
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 01:38:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 16 01:19:03 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.933 min
Delta R.T.: 0.000 min
Response: 432384
Conc: 4.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.938 min
Delta R.T.: 0.000 min
Response: 144015
Conc: 4.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.000 min
Delta R.T.: 0.000 min
Response: 338420
Conc: 4.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.232 min
Delta R.T.: 0.000 min
Response: 140242
Conc: 4.99 ng/ml



#3 AR-1016-1


R.T.: 6.265 min
Delta R.T.: 0.000 min
Response: 161655
Conc: 50.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 5.191 min
Delta R.T.: 0.000 min
Response: 66519
Conc: 54.06 ng/ml




#4 AR-1016-2

R.T.: 6.288 min
Delta R.T.: 0.000 min
Response: 233206
Conc: 51.24 ng/ml



#4 AR-1016-2


R.T.: 5.213 min
Delta R.T.: 0.000 min
Response: 94241
Conc: 54.88 ng/ml



#5 AR-1016-3


R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 134545
Conc: 49.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 5.402 min
Delta R.T.: 0.000 min
Response: 47360
Conc: 52.03 ng/ml




#6 AR-1016-4

R.T.: 6.462 min
Delta R.T.: 0.000 min
Response: 111893
Conc: 50.39 ng/ml



#6 AR-1016-4


R.T.: 5.456 min
Delta R.T.: 0.000 min
Response: 35393
Conc: 47.91 ng/ml



#7 AR-1016-5


R.T.: 6.778 min
Delta R.T.: 0.000 min
Response: 103231
Conc: 48.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 46104
Conc: 50.85 ng/ml




#31 AR-1260-1

R.T.: 7.961 min
Delta R.T.: 0.000 min
Response: 157332
Conc: 48.56 ng/ml



#31 AR-1260-1


R.T.: 6.780 min
Delta R.T.: 0.000 min
Response: 99928
Conc: 59.62 ng/ml



#32 AR-1260-2


R.T.: 8.229 min
Delta R.T.: 0.000 min
Response: 193293
Conc: 49.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#32 AR-1260-2

R.T.: 6.979 min
Delta R.T.: 0.000 min
Response: 109735
Conc: 53.15 ng/ml




#33 AR-1260-3

R.T.: 8.594 min
Delta R.T.: 0.000 min
Response: 152490
Conc: 51.48 ng/ml



#33 AR-1260-3


R.T.: 7.132 min
Delta R.T.: 0.000 min
Response: 92155
Conc: 52.06 ng/ml



#34 AR-1260-4


R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 174433
Conc: 50.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#34 AR-1260-4

R.T.: 7.616 min
Delta R.T.: 0.000 min
Response: 69469
Conc: 47.15 ng/ml



#35 AR-1260-5

R.T.: 9.157 min
Delta R.T.: 0.000 min
Response: 340629
Conc: 50.56 ng/ml



#35 AR-1260-5

R.T.: 7.865 min
Delta R.T.: 0.000 min
Response: 184071
Conc: 47.56 ng/ml