

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121523\
 Data File : PO100526.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2023 18:05
 Operator : YP/AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 04:48:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 16 04:47:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.467	3.695	83168696	77273786	50.000	50.000
2) SA Decachlor...	10.327	8.773	102.0E6	113.5E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.773	7.643	79678426	90781856	500.000	500.000
42) L9 AR-1268-2	8.869	7.709	69532739	83096776	500.000	500.000
43) L9 AR-1268-3	9.105	7.917	64789527	77142342	500.000	500.000
44) L9 AR-1268-4	9.545	8.209	22757871	27611060	500.000	500.000
45) L9 AR-1268-5	9.975	8.509	187.5E6	215.0E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO121523\
Data File : PO100526.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2023 18:05
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 04:48:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO121523.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 16 04:47:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

