

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121523\
 Data File : PO100524.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2023 17:32
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 05:05:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 16 05:05:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.469	3.695	164.0E6	151.0E6	99.304	98.826
2) SA Decachlor...	10.330	8.773	194.3E6	217.3E6	97.567	97.810
Target Compounds						
41) L9 AR-1268-1	8.774	7.644	152.8E6	176.0E6	979.113	984.421
42) L9 AR-1268-2	8.871	7.709	132.9E6	161.9E6	977.453	986.919
43) L9 AR-1268-3	9.106	7.917	125.0E6	149.4E6	981.843	983.919
44) L9 AR-1268-4	9.546	8.210	43951957	52164136	982.521	971.524
45) L9 AR-1268-5	9.977	8.509	364.6E6	424.3E6	985.833	993.338



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121523\
Data File : PO100524.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2023 17:32
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 05:05:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121523.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 16 05:05:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.469 min
Delta R.T.: 0.000 min
Response: 164038014
Conc: 99.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.695 min
Delta R.T.: 0.000 min
Response: 150961468
Conc: 98.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.330 min
Delta R.T.: 0.000 min
Response: 194325906
Conc: 97.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.773 min
Delta R.T.: 0.000 min
Response: 217272425
Conc: 97.81 ng/ml



#41 AR-1268-1


R.T.: 8.774 min
Delta R.T.: 0.000 min
Response: 152836163
Conc: 979.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.644 min
Delta R.T.: 0.000 min
Response: 175993469
Conc: 984.42 ng/ml




#42 AR-1268-2

R.T.: 8.871 min
Delta R.T.: 0.000 min
Response: 132932674
Conc: 977.45 ng/ml



#42 AR-1268-2


R.T.: 7.709 min
Delta R.T.: 0.000 min
Response: 161901893
Conc: 986.92 ng/ml



#43 AR-1268-3


R.T.: 9.106 min
Delta R.T.: 0.000 min
Response: 124957399
Conc: 981.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 7.917 min
Delta R.T.: 0.000 min
Response: 149401073
Conc: 983.92 ng/ml




#44 AR-1268-4

R.T.: 9.546 min
Delta R.T.: 0.000 min
Response: 43951957
Conc: 982.52 ng/ml



#44 AR-1268-4


R.T.: 8.210 min
Delta R.T.: 0.000 min
Response: 52164136
Conc: 971.52 ng/ml



#45 AR-1268-5

R.T.: 9.977 min
Delta R.T.: 0.000 min
Response: 364560605
Conc: 985.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.509 min
Delta R.T.: 0.000 min
Response: 424260957
Conc: 993.34 ng/ml