

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121523\
 Data File : PO100532.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2023 19:47
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO121523

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 05:23:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 16 05:19:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.467	3.695	83006016	79191132	52.474	52.151
2)	SA Decachlor...	10.326	8.773	57892411	65895761	51.560	52.914
Target Compounds							
26)	L6 AR-1254-1	6.518	5.608	28405678	25945245	518.070	515.850
27)	L6 AR-1254-2	6.738	5.756	41662791	23377609	516.013	508.222
28)	L6 AR-1254-3	7.109	6.165	40291255	39165053	527.038	518.140
29)	L6 AR-1254-4	7.397	6.395	22418109	21460736	516.944	508.290
30)	L6 AR-1254-5	7.821	6.817	28494519	37697297	514.079	514.151



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121523\
Data File : PO100532.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2023 19:47
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO121523

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 05:23:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121523.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 16 05:19:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.467 min
Delta R.T.: 0.000 min
Response: 83006016
Conc: 52.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121523




#1 Tetrachloro-m-xylene

R.T.: 3.695 min
Delta R.T.: 0.000 min
Response: 79191132
Conc: 52.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.326 min
Delta R.T.: 0.000 min
Response: 57892411
Conc: 51.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.773 min
Delta R.T.: 0.000 min
Response: 65895761
Conc: 52.91 ng/ml



#26 AR-1254-1


R.T.: 6.518 min
Delta R.T.: 0.000 min
Response: 28405678
Conc: 518.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121523




#26 AR-1254-1

R.T.: 5.608 min
Delta R.T.: 0.000 min
Response: 25945245
Conc: 515.85 ng/ml




#27 AR-1254-2

R.T.: 6.738 min
Delta R.T.: -0.001 min
Response: 41662791
Conc: 516.01 ng/ml



#27 AR-1254-2


R.T.: 5.756 min
Delta R.T.: 0.000 min
Response: 23377609
Conc: 508.22 ng/ml



#28 AR-1254-3


R.T.: 7.109 min
Delta R.T.: 0.000 min
Response: 40291255
Conc: 527.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121523




#28 AR-1254-3

R.T.: 6.165 min
Delta R.T.: -0.001 min
Response: 39165053
Conc: 518.14 ng/ml




#29 AR-1254-4

R.T.: 7.397 min
Delta R.T.: 0.000 min
Response: 22418109
Conc: 516.94 ng/ml



#29 AR-1254-4


R.T.: 6.395 min
Delta R.T.: 0.000 min
Response: 21460736
Conc: 508.29 ng/ml



#30 AR-1254-5

R.T.: 7.821 min
Delta R.T.: 0.000 min
Response: 28494519
Conc: 514.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO121523



#30 AR-1254-5

R.T.: 6.817 min
Delta R.T.: 0.000 min
Response: 37697297
Conc: 514.15 ng/ml