

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121621\
 Data File : P0083779.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Dec 2021 22:53
 Operator : AJ\MA
 Sample : M5070-12 10X
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CLI-2021-14-A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 17 06:03:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 16 04:05:27 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.937	3.942	114380	43096	1.201	1.469
2) SA Decachlor...	11.016	9.238	114391	45108	1.670	1.606
Target Compounds						
30) L6 AR-1254-5	8.517	0.000	4754	0	1.334	N.D. #
33) L7 AR-1260-3	8.585	0.000	17217	0	5.812	N.D. #
34) L7 AR-1260-4	8.830	0.000	7801	0	2.281	N.D. #
35) L7 AR-1260-5	9.165	0.000	7624	0	1.132	N.D. #
36) L8 AR-1262-1	8.585	0.000	17217	0	4.269	N.D. #
37) L8 AR-1262-2	9.165	0.000	7624	0	1.090	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121621\
Data File : P0083779.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Dec 2021 22:53
Operator : AJ\MA
Sample : M5070-12 10X
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CLI-2021-14-A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 17 06:03:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 16 04:05:27 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.937 min
Delta R.T.: 0.004 min
Response: 114380
Conc: 1.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
CLI-2021-14-A




#1 Tetrachloro-m-xylene

R.T.: 3.942 min
Delta R.T.: 0.003 min
Response: 43096
Conc: 1.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.016 min
Delta R.T.: 0.011 min
Response: 114391
Conc: 1.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.238 min
Delta R.T.: 0.007 min
Response: 45108
Conc: 1.61 ng/ml



#30 AR-1254-5


R.T.: 8.517 min
Delta R.T.: -0.006 min
Response: 4754
Conc: 1.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
CLI-2021-14-A




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.313 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.585 min
Delta R.T.: -0.011 min
Response: 17217
Conc: 5.81 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.133 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 8.830 min
Delta R.T.: 0.002 min
Response: 7801
Conc: 2.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
CLI-2021-14-A




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.616 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 9.165 min
Delta R.T.: 0.008 min
Response: 7624
Conc: 1.13 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T. : 7.866 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 8.585 min
Delta R.T.: -0.012 min
Response: 17217
Conc: 4.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
CLI-2021-14-A




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.312 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 9.165 min
Delta R.T.: 0.006 min
Response: 7624
Conc: 1.09 ng/ml



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.866 min
Response: 0
Conc: N.D.