

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121622\
 Data File : PO091444.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Dec 2022 14:43
 Operator : YP/AJ
 Sample : N6071-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 FVSUB-TP4-COMP-04

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/19/2022
 Supervised By :Ankita Jodhani 12/19/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 20:49:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.298	3.473	95075760	33114516	28.120	31.551
2) SA Decachlor...	10.054	8.479	62656474	24019736	23.589	24.740
Target Compounds						
31) L7 AR-1260-1	7.091	6.047	10711035	4308966	69.821	74.418
32) L7 AR-1260-2	7.349	6.237	12862161	4480201	73.690	65.609
33) L7 AR-1260-3	7.713	6.390	11681079	4565856	102.760	69.609 #
34) L7 AR-1260-4	7.938	6.864	14671487	4495383	108.270m	94.124
35) L7 AR-1260-5	8.256	7.109	27902676	11071612	109.852	101.843

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121622\
Data File : P0091444.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Dec 2022 14:43
Operator : YP/AJ
Sample : N6071-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
FVSUB-TP4-COMP-04

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/19/2022
Supervised By :Ankita Jodhani 12/19/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 20:49:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

