

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121622\
 Data File : PO091468.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2022 01:40
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 02:11:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.296	3.473	167.5E6	56994848	49.547	54.304
2) SA Decachlor...	10.048	8.476	120.7E6	50693777	45.432	52.215
Target Compounds						
21) L5 AR-1248-1	5.472	4.557	34120176	11932617	478.993	507.214
22) L5 AR-1248-2	5.749	4.795	46194953	16722260	435.764	510.761
23) L5 AR-1248-3	5.953	4.836	52442219	17394646	445.629	512.514
24) L5 AR-1248-4	6.361	5.007	55557734	20204589	454.319	506.351
25) L5 AR-1248-5	6.399	5.402	54609062	18758019	476.118	523.309



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121622\
Data File : PO091468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2022 01:40
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 02:11:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.296 min
Delta R.T.: -0.002 min
Response: 167524805
Conc: 49.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.473 min
Delta R.T.: -0.003 min
Response: 56994848
Conc: 54.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.048 min
Delta R.T.: -0.002 min
Response: 120675770
Conc: 45.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.476 min
Delta R.T.: -0.006 min
Response: 50693777
Conc: 52.21 ng/ml



#21 AR-1248-1


R.T.: 5.472 min
Delta R.T.: -0.003 min
Response: 34120176
Conc: 478.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.557 min
Delta R.T.: -0.004 min
Response: 11932617
Conc: 507.21 ng/ml




#22 AR-1248-2

R.T.: 5.749 min
Delta R.T.: -0.003 min
Response: 46194953
Conc: 435.76 ng/ml



#22 AR-1248-2


R.T.: 4.795 min
Delta R.T.: -0.005 min
Response: 16722260
Conc: 510.76 ng/ml



#23 AR-1248-3


R.T.: 5.953 min
Delta R.T.: -0.003 min
Response: 52442219
Conc: 445.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.836 min
Delta R.T.: -0.004 min
Response: 17394646
Conc: 512.51 ng/ml




#24 AR-1248-4

R.T.: 6.361 min
Delta R.T.: -0.002 min
Response: 55557734
Conc: 454.32 ng/ml



#24 AR-1248-4


R.T.: 5.007 min
Delta R.T.: -0.005 min
Response: 20204589
Conc: 506.35 ng/ml



#25 AR-1248-5

R.T.: 6.399 min
Delta R.T.: -0.003 min
Response: 54609062
Conc: 476.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.402 min
Delta R.T.: -0.005 min
Response: 18758019
Conc: 523.31 ng/ml