

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121918\
 Data File : P0052790.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Dec 2018 9:24
 Operator : SM/SJ
 Sample : J6446-20
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P001-SS023-1218-01

Manual Integrations
 APPROVED

Sohil
 1/7/2019 8:33:33 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 10:29:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Dec 16 03:01:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.245	3.270	15180939	3890172	19.358	18.358
2)	SA Decachlor...	9.787	7.951	4915511	1951911	9.756m	8.847
Target Compounds							
31)	L7 AR-1260-1	6.981	5.652	48641643	19201397	1665.036	1400.143
32)	L7 AR-1260-2	7.237	5.836	68594182	27186580	1961.067	1636.029
33)	L7 AR-1260-3	7.592	5.977	47100100	20165295	2023.714	1318.919 #
34)	L7 AR-1260-4	7.817	6.427	49981929	17403584	2007.557	1729.604
35)	L7 AR-1260-5	8.125	6.669	95183838	39061703	1941.407	1647.951

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121918\
Data File : PO052790.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Dec 2018 9:24
Operator : SM/SJ
Sample : J6446-20
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
P001-SS023-1218-01

Manual Integrations
APPROVED

Sohil
1/7/2019 8:33:33 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 10:29:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121518.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Dec 16 03:01:24 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

