Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122019\
Data File : PO064820.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Dec 2019 12:27
Operator : SM/AJ
Sample : K6358-05DL 1000X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId : PCB-121619-05DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 13:11:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
	Target Compounds						
26)	L6 AR-1254-1	6.181	5.313	2184176	4059748	1715.048	1999.161
27)	L6 AR-1254-2	6.399	5.459	3589150	3506096	1725.513	1922.852
28)	L6 AR-1254-3	6.761	5.858	4590052	7084576	1922.300	2280.619
29)	L6 AR-1254-4	7.044	6.085	3576540	4525223	1912.651	2203.093
30)	L6 AR-1254-5	7.460	6.499	4696903	7799243	2232.343	2612.913

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122019\
Data File : PO064820.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Dec 2019 12:27
Operator : SM/AJ
Sample : K6358-05DL 1000X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId : PCB-121619-05DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 13:11:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.177 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
PCB-121619-05DL



#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.457 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.688 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.376 min
Response: 0
Conc: N.D.



#26 AR-1254-1

R.T.: 6.181 min
Delta R.T.: 0.011 min
Response: 2184176
Conc: 1715.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-121619-05DL



#26 AR-1254-1

R.T.: 5.313 min
Delta R.T.: 0.000 min
Response: 4059748
Conc: 1999.16 ng/ml



#27 AR-1254-2

R.T.: 6.399 min
Delta R.T.: 0.012 min
Response: 3589150
Conc: 1725.51 ng/ml



#27 AR-1254-2

R.T.: 5.459 min
Delta R.T.: 0.000 min
Response: 3506096
Conc: 1922.85 ng/ml



#28 AR-1254-3

R.T.: 6.761 min
Delta R.T.: 0.013 min
Response: 4590052
Conc: 1922.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-121619-05DL



#28 AR-1254-3

R.T.: 5.858 min
Delta R.T.: 0.000 min
Response: 7084576
Conc: 2280.62 ng/ml



#29 AR-1254-4

R.T.: 7.044 min
Delta R.T.: 0.012 min
Response: 3576540
Conc: 1912.65 ng/ml



#29 AR-1254-4

R.T.: 6.085 min
Delta R.T.: 0.000 min
Response: 4525223
Conc: 2203.09 ng/ml



#30 AR-1254-5

R.T.: 7.460 min
Delta R.T.: 0.013 min
Response: 4696903
Conc: 2232.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-121619-05DL



#30 AR-1254-5

R.T.: 6.499 min
Delta R.T.: 0.000 min
Response: 7799243
Conc: 2612.91 ng/ml