

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122019\
 Data File : P0064842.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Dec 2019 18:41
 Operator : SM/AJ
 Sample : K6315-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 NCQ-632

Manual Integrations
 APPROVED

Ankita
 12/23/2019 11:09:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 05:28:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.183	3.458	564923	666633	23.205m	26.298m
2) SA Decachlor...	9.698	8.373	481089	613255	15.676	17.659

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122019\
Data File : P0064842.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Dec 2019 18:41
Operator : SM/AJ
Sample : K6315-03
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
NCQ-632


Manual Integrations
APPROVED

Ankita
12/23/2019 11:09:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 05:28:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.183 min
Delta R.T.: 0.006 min
Response: 564923
Conc: 23.20 ng/ml m

Instrument :
ECD_O
ClientSampleId :
NCQ-632


Manual Integrations
APPROVED

Ankita
12/23/2019 11:09:58 AM




#1 Tetrachloro-m-xylene

R.T.: 3.458 min
Delta R.T.: 0.000 min
Response: 666633
Conc: 26.30 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.698 min
Delta R.T.: 0.011 min
Response: 481089
Conc: 15.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.373 min
Delta R.T.: -0.003 min
Response: 613255
Conc: 17.66 ng/ml