

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122021\
 Data File : PO083860.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Dec 2021 12:34
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 13:10:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 16 04:05:27 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.934	3.941	4992586	1555015	52.405	53.012
2) SA Decachlor...	11.009	9.236	3703328	1526407	54.068	54.361
Target Compounds						
26) L6 AR-1254-1	7.182	6.066	1559249	915456	506.222	498.893
27) L6 AR-1254-2	7.413	6.226	2368406	801025	510.358	504.343
28) L6 AR-1254-3	7.798	6.646	2391191	1231978	512.813	514.334
29) L6 AR-1254-4	8.096	6.887	1754446	832849	513.250	495.685
30) L6 AR-1254-5	8.527	7.315	1855033	1149589	520.342	541.419



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122021\
Data File : P0083860.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Dec 2021 12:34
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 13:10:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 16 04:05:27 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.934 min
Delta R.T.: 0.001 min
Response: 4992586
Conc: 52.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.941 min
Delta R.T.: 0.002 min
Response: 1555015
Conc: 53.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.009 min
Delta R.T.: 0.004 min
Response: 3703328
Conc: 54.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.236 min
Delta R.T.: 0.005 min
Response: 1526407
Conc: 54.36 ng/ml



#26 AR-1254-1


R.T.: 7.182 min
Delta R.T.: 0.002 min
Response: 1559249
Conc: 506.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.066 min
Delta R.T.: 0.003 min
Response: 915456
Conc: 498.89 ng/ml




#27 AR-1254-2

R.T.: 7.413 min
Delta R.T.: 0.003 min
Response: 2368406
Conc: 510.36 ng/ml



#27 AR-1254-2


R.T.: 6.226 min
Delta R.T.: 0.002 min
Response: 801025
Conc: 504.34 ng/ml



#28 AR-1254-3


R.T.: 7.798 min
Delta R.T.: 0.003 min
Response: 2391191
Conc: 512.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.646 min
Delta R.T.: 0.003 min
Response: 1231978
Conc: 514.33 ng/ml




#29 AR-1254-4

R.T.: 8.096 min
Delta R.T.: 0.004 min
Response: 1754446
Conc: 513.25 ng/ml



#29 AR-1254-4


R.T.: 6.887 min
Delta R.T.: 0.003 min
Response: 832849
Conc: 495.68 ng/ml



#30 AR-1254-5

R.T.: 8.527 min
Delta R.T.: 0.003 min
Response: 1855033
Conc: 520.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.315 min
Delta R.T.: 0.002 min
Response: 1149589
Conc: 541.42 ng/ml