

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122022\
 Data File : PO091554.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2022 20:54
 Operator : YP/AJ
 Sample : N6111-01 10X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC274770-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/21/2022
 Supervised By :Ankita Jodhani 12/21/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 08:45:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.293	3.470	5825152	1909720	1.723	1.820
2) SA Decachlor...	10.032	8.466	5592901	2181928	2.106	2.247
Target Compounds						
26) L6 AR-1254-1	6.330	5.356	1064935	328167	8.479m	5.599 #
27) L6 AR-1254-2	6.550	5.505	1280439	379585	6.791	7.112
28) L6 AR-1254-3	6.917	5.908	1250585	535593	6.672	6.574
29) L6 AR-1254-4	7.204	6.138	962764	247713	6.969	5.063 #
30) L6 AR-1254-5	7.627	6.557	755465	276185	4.924	3.705

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122022\
Data File : PO091554.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2022 20:54
Operator : YP/AJ
Sample : N6111-01 10X
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC274770-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/21/2022
Supervised By :Ankita Jodhani 12/21/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 08:45:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.293 min
Delta R.T.: -0.005 min
Response: 5825152
Conc: 1.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC274770-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/21/2022
Supervised By :Ankita Jodhani 12/21/2022




#1 Tetrachloro-m-xylene

R.T.: 3.470 min
Delta R.T.: -0.007 min
Response: 1909720
Conc: 1.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.032 min
Delta R.T.: -0.017 min
Response: 5592901
Conc: 2.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.466 min
Delta R.T.: -0.016 min
Response: 2181928
Conc: 2.25 ng/ml




#26 AR-1254-1

R.T.: 6.330 min
Delta R.T.: -0.008 min
Response: 1064935
Conc: 8.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC274770-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/21/2022
Supervised By :Ankita Jodhani 12/21/2022




#26 AR-1254-1

R.T.: 5.356 min
Delta R.T.: -0.012 min
Response: 328167
Conc: 5.60 ng/ml




#27 AR-1254-2

R.T.: 6.550 min
Delta R.T.: -0.010 min
Response: 1280439
Conc: 6.79 ng/ml



#27 AR-1254-2

R.T.: 5.505 min
Delta R.T.: -0.012 min
Response: 379585
Conc: 7.11 ng/ml




#28 AR-1254-3

R.T.: 6.917 min
Delta R.T.: -0.010 min
Response: 1250585
Conc: 6.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC274770-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/21/2022
Supervised By :Ankita Jodhani 12/21/2022




#28 AR-1254-3

R.T.: 5.908 min
Delta R.T.: -0.013 min
Response: 535593
Conc: 6.57 ng/ml




#29 AR-1254-4

R.T.: 7.204 min
Delta R.T.: -0.012 min
Response: 962764
Conc: 6.97 ng/ml



#29 AR-1254-4

R.T.: 6.138 min
Delta R.T.: -0.014 min
Response: 247713
Conc: 5.06 ng/ml




#30 AR-1254-5

R.T.: 7.627 min
Delta R.T.: -0.011 min
Response: 755465
Conc: 4.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC274770-1-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/21/2022
Supervised By :Ankita Jodhani 12/21/2022



#30 AR-1254-5

R.T.: 6.557 min
Delta R.T.: -0.015 min
Response: 276185
Conc: 3.70 ng/ml