

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122023\
 Data File : PO100602.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2023 00:39
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 03:25:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 03:19:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.349	3.646	247.2E6	159.0E6	75.364	75.337
2)	SA Decachlor...	10.091	8.706	160.6E6	116.1E6	75.153	75.132
Target Compounds							
26)	L6 AR-1254-1	6.374	5.552	75270939	74146952	746.735	752.032
27)	L6 AR-1254-2	6.593	5.701	109.6E6	65457186	752.106	750.963
28)	L6 AR-1254-3	6.960	6.109	106.5E6	99343751	755.633	750.962
29)	L6 AR-1254-4	7.247	6.339	67002781	53142026	755.973	752.750
30)	L6 AR-1254-5	7.668	6.760	84357135	82634958	753.294	751.072


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122023\
Data File : PO100602.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2023 00:39
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 03:25:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 03:19:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.349 min
Delta R.T.: 0.000 min
Response: 247235648
Conc: 75.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.000 min
Response: 159028532
Conc: 75.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.091 min
Delta R.T.: 0.000 min
Response: 160649925
Conc: 75.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.706 min
Delta R.T.: 0.000 min
Response: 116057764
Conc: 75.13 ng/ml



#26 AR-1254-1


R.T.: 6.374 min
Delta R.T.: 0.000 min
Response: 75270939
Conc: 746.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.552 min
Delta R.T.: 0.000 min
Response: 74146952
Conc: 752.03 ng/ml




#27 AR-1254-2

R.T.: 6.593 min
Delta R.T.: 0.000 min
Response: 109646594
Conc: 752.11 ng/ml



#27 AR-1254-2


R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 65457186
Conc: 750.96 ng/ml



#28 AR-1254-3


R.T.: 6.960 min
Delta R.T.: 0.000 min
Response: 106453173
Conc: 755.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 99343751
Conc: 750.96 ng/ml




#29 AR-1254-4

R.T.: 7.247 min
Delta R.T.: 0.000 min
Response: 67002781
Conc: 755.97 ng/ml



#29 AR-1254-4


R.T.: 6.339 min
Delta R.T.: 0.000 min
Response: 53142026
Conc: 752.75 ng/ml



#30 AR-1254-5

R.T.: 7.668 min
Delta R.T.: 0.000 min
Response: 84357135
Conc: 753.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 82634958
Conc: 751.07 ng/ml