

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122023\
 Data File : PO100616.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2023 04:33
 Operator : YP/AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO122023AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 05:20:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.350	3.646	171.7E6	112.0E6	55.043	54.828
2) SA Decachlor...	10.091	8.705	188.8E6	136.6E6	90.364	90.519
Target Compounds						
41) L9 AR-1268-1	8.595	7.586	146.3E6	116.6E6	517.845	525.943
42) L9 AR-1268-2	8.689	7.650	130.3E6	105.7E6	515.645	522.836
43) L9 AR-1268-3	8.915	7.858	118.9E6	96798727	521.212	513.071
44) L9 AR-1268-4	9.340	8.148	41148474	34713197	525.513	517.013
45) L9 AR-1268-5	9.753	8.445	374.2E6	268.3E6	521.855	517.070



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122023\
Data File : PO100616.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2023 04:33
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO122023AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 05:20:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.350 min
Delta R.T.: 0.000 min
Response: 171748400
Conc: 55.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO122023AR1268




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.000 min
Response: 112037802
Conc: 54.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.091 min
Delta R.T.: -0.001 min
Response: 188781404
Conc: 90.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.705 min
Delta R.T.: 0.000 min
Response: 136625874
Conc: 90.52 ng/ml



#41 AR-1268-1


R.T.: 8.595 min
Delta R.T.: 0.000 min
Response: 146322437
Conc: 517.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO122023AR1268




#41 AR-1268-1

R.T.: 7.586 min
Delta R.T.: 0.000 min
Response: 116580476
Conc: 525.94 ng/ml




#42 AR-1268-2

R.T.: 8.689 min
Delta R.T.: 0.000 min
Response: 130312236
Conc: 515.65 ng/ml



#42 AR-1268-2


R.T.: 7.650 min
Delta R.T.: 0.000 min
Response: 105748705
Conc: 522.84 ng/ml



#43 AR-1268-3


R.T.: 8.915 min
Delta R.T.: 0.000 min
Response: 118922311
Conc: 521.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO122023AR1268




#43 AR-1268-3

R.T.: 7.858 min
Delta R.T.: 0.000 min
Response: 96798727
Conc: 513.07 ng/ml




#44 AR-1268-4

R.T.: 9.340 min
Delta R.T.: 0.000 min
Response: 41148474
Conc: 525.51 ng/ml



#44 AR-1268-4


R.T.: 8.148 min
Delta R.T.: -0.001 min
Response: 34713197
Conc: 517.01 ng/ml



#45 AR-1268-5

R.T.: 9.753 min
Delta R.T.: -0.001 min
Response: 374183490
Conc: 521.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO122023AR1268



#45 AR-1268-5

R.T.: 8.445 min
Delta R.T.: 0.000 min
Response: 268268516
Conc: 517.07 ng/ml