

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO122024\
 Data File : PO108728.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2024 21:42
 Operator : YP/AJ
 Sample : P5341-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 STORMWATER-AQ-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 01:26:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 05:58:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.707	3.704	158.7E6	96224520	18.243	18.982
2)	SA Decachlor...	8.774	8.724	77374624	45700697	10.593	11.794
Target Compounds							
3)	L1 AR-1016-1	0.000	4.809	0	541870	N.D.	3.377 #
4)	L1 AR-1016-2	0.000	4.827	0	798531	N.D.	3.595 #
5)	L1 AR-1016-3	0.000	4.996	0	60874	N.D.	0.484 #
6)	L1 AR-1016-4	0.000	5.042	0	318884	N.D.	3.043 #
7)	L1 AR-1016-5	0.000	5.249	0	2861633	N.D.	21.164 #
8)	L2 AR-1221-1	3.907	3.903	11264680	6158881	99.561	99.141
9)	L2 AR-1221-2	4.009	4.004	13100783	6508775	151.069	137.186
10)	L2 AR-1221-3	4.085	4.078	13704865	3355616	53.913	23.869 #
11)	L3 AR-1232-1	4.085	4.078	13704865	3355616	66.928	29.947 #
12)	L3 AR-1232-2	0.000	4.809	0	541870	N.D.	5.124 #
13)	L3 AR-1232-3	0.000	4.996	0	60874	N.D.	1.009 #
14)	L3 AR-1232-4	0.000	5.071	0	204954	N.D.	3.689 #
15)	L3 AR-1232-5	0.000	5.249	0	2861633	N.D.	49.702 #
16)	L4 AR-1242-1	0.000	4.809	0	541870	N.D.	3.991 #
17)	L4 AR-1242-2	0.000	4.809	0	541870	N.D.	2.922 #
18)	L4 AR-1242-3	0.000	4.996	0	60874	N.D.	0.578 #
19)	L4 AR-1242-4	0.000	5.071	0	204954	N.D.	1.910 #
20)	L4 AR-1242-5	5.698f	5.598	20424926	2701318	98.352	21.371 #
21)	L5 AR-1248-1	0.000	4.809	0	541870	N.D.	5.155 #
22)	L5 AR-1248-2	0.000	5.042	0	318884	N.D.	2.135 #
23)	L5 AR-1248-3	0.000	5.071	0	204954	N.D.	1.294 #
24)	L5 AR-1248-4	5.619	5.249	7934336	2861633	16.650	15.676
25)	L5 AR-1248-5	5.698f	5.637	20424926	1773607	60.320	10.100 #
26)	L6 AR-1254-1	5.619	5.598	7934336	2701318	15.447	9.974 #
27)	L6 AR-1254-2	0.000	5.740	0	2033653	N.D.	8.473 #
28)	L6 AR-1254-3	6.169	6.160	15709063	1250783	21.814	3.331 #
29)	L6 AR-1254-4	6.434	6.366	4503178	2311685	10.262	10.620
30)	L6 AR-1254-5	6.818	6.789	10058519	7162277	15.837	22.191 #
31)	L7 AR-1260-1	6.300	6.275	19227446	6911716	42.060	29.533 #
32)	L7 AR-1260-2	6.489	6.462	24656349	11166798	44.368	39.801
33)	L7 AR-1260-3	6.858	6.617	6397701	9538893	13.798	36.191 #
34)	L7 AR-1260-4	7.118	7.088	3620925	4277700	8.513	20.107 #
35)	L7 AR-1260-5	7.360	7.328	9490797	7568688	9.765	15.640 #
36)	L8 AR-1262-1	6.858	6.829	6397701	5749460	9.717	17.012 #
37)	L8 AR-1262-2	7.360	7.328	9490797	7568688	8.537	13.235 #
38)	L8 AR-1262-3	7.646	7.611	1328687	963998	3.041	4.295 #
39)	L8 AR-1262-4	7.709	7.674	8103390	4144136	10.050	10.118
40)	L8 AR-1262-5	8.208	8.168	-2869190	2473091	N.D.	13.380 #
41)	L9 AR-1268-1	7.646	7.611	1328687	963998	1.068	1.492 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122024\
 Data File : PO108728.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2024 21:42
 Operator : YP/AJ
 Sample : P5341-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 STORMWATER-AQ-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 01:26:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 05:58:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.709	7.674	8103390	4144136	7.224	7.126
43) L9	AR-1268-3	0.000	7.884	0	403923	N.D.	0.810 #
44) L9	AR-1268-4	8.208	8.168	-2869190	2473091	N.D.	12.180 #
45) L9	AR-1268-5	8.509	8.464	2255232	1046832	0.814	0.705



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122024\
Data File : PO108728.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2024 21:42
Operator : YP/AJ
Sample : P5341-01
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 01:26:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 05:58:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.707 min
Delta R.T.: -0.003 min
Response: 158719860
Conc: 18.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#1 Tetrachloro-m-xylene

R.T.: 3.704 min
Delta R.T.: -0.005 min
Response: 96224520
Conc: 18.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.774 min
Delta R.T.: -0.017 min
Response: 77374624
Conc: 10.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.724 min
Delta R.T.: -0.017 min
Response: 45700697
Conc: 11.79 ng/ml



#3 AR-1016-1


R.T.: 0.000 min
Exp R.T.: 4.810 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#3 AR-1016-1

R.T.: 4.809 min
Delta R.T.: 0.011 min
Response: 541870
Conc: 3.38 ng/ml




#4 AR-1016-2

R.T.: 0.000 min
Exp R.T.: 4.830 min
Response: 0
Conc: N.D.



#4 AR-1016-2


R.T.: 4.827 min
Delta R.T.: 0.009 min
Response: 798531
Conc: 3.60 ng/ml



#5 AR-1016-3


R.T.: 0.000 min
Exp R.T.: 4.885 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#5 AR-1016-3

R.T.: 4.996 min
Delta R.T.: 0.002 min
Response: 60874
Conc: 0.48 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T.: 5.007 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 5.042 min
Delta R.T.: 0.006 min
Response: 318884
Conc: 3.04 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 5.265 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#7 AR-1016-5

R.T.: 5.249 min
Delta R.T.: -0.001 min
Response: 2861633
Conc: 21.16 ng/ml




#8 AR-1221-1

R.T.: 3.907 min
Delta R.T.: -0.019 min
Response: 11264680
Conc: 99.56 ng/ml



#8 AR-1221-1


R.T.: 3.903 min
Delta R.T.: -0.019 min
Response: 6158881
Conc: 99.14 ng/ml



#9 AR-1221-2


R.T.: 4.009 min
Delta R.T.: -0.002 min
Response: 13100783
Conc: 151.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#9 AR-1221-2

R.T.: 4.004 min
Delta R.T.: -0.003 min
Response: 6508775
Conc: 137.19 ng/ml




#10 AR-1221-3

R.T.: 4.085 min
Delta R.T.: -0.004 min
Response: 13704865
Conc: 53.91 ng/ml



#10 AR-1221-3


R.T.: 4.078 min
Delta R.T.: -0.005 min
Response: 3355616
Conc: 23.87 ng/ml



#11 AR-1232-1


R.T.: 4.085 min
Delta R.T.: -0.004 min
Response: 13704865
Conc: 66.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#11 AR-1232-1

R.T.: 4.078 min
Delta R.T.: -0.006 min
Response: 3355616
Conc: 29.95 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 4.586 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 4.809 min
Delta R.T.: -0.008 min
Response: 541870
Conc: 5.12 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T.: 4.829 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#13 AR-1232-3

R.T.: 4.996 min
Delta R.T.: 0.002 min
Response: 60874
Conc: 1.01 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T.: 5.005 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 5.071 min
Delta R.T.: -0.007 min
Response: 204954
Conc: 3.69 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 5.048 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#15 AR-1232-5

R.T.: 5.249 min
Delta R.T.: 0.000 min
Response: 2861633
Conc: 49.70 ng/ml




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T.: 4.808 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 4.809 min
Delta R.T.: 0.012 min
Response: 541870
Conc: 3.99 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T.: 4.828 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#17 AR-1242-2

R.T.: 4.809 min
Delta R.T.: -0.008 min
Response: 541870
Conc: 2.92 ng/ml




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T.: 4.884 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 4.996 min
Delta R.T.: 0.002 min
Response: 60874
Conc: 0.58 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 5.005 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#19 AR-1242-4

R.T.: 5.071 min
Delta R.T.: -0.006 min
Response: 204954
Conc: 1.91 ng/ml




#20 AR-1242-5

R.T.: 5.698 min
Delta R.T.: 0.038 min
Response: 20424926
Conc: 98.35 ng/ml



#20 AR-1242-5


R.T.: 5.598 min
Delta R.T.: -0.003 min
Response: 2701318
Conc: 21.37 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 4.807 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#21 AR-1248-1

R.T.: 4.809 min
Delta R.T.: 0.012 min
Response: 541870
Conc: 5.15 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T.: 5.047 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 5.042 min
Delta R.T.: 0.007 min
Response: 318884
Conc: 2.14 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 5.263 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#23 AR-1248-3

R.T.: 5.071 min
Delta R.T.: -0.006 min
Response: 204954
Conc: 1.29 ng/ml




#24 AR-1248-4

R.T.: 5.619 min
Delta R.T.: 0.001 min
Response: 7934336
Conc: 16.65 ng/ml



#24 AR-1248-4


R.T.: 5.249 min
Delta R.T.: 0.000 min
Response: 2861633
Conc: 15.68 ng/ml



#25 AR-1248-5


R.T.: 5.698 min
Delta R.T.: 0.038 min
Response: 20424926
Conc: 60.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#25 AR-1248-5

R.T.: 5.637 min
Delta R.T.: -0.005 min
Response: 1773607
Conc: 10.10 ng/ml




#26 AR-1254-1

R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 7934336
Conc: 15.45 ng/ml



#26 AR-1254-1


R.T.: 5.598 min
Delta R.T.: -0.004 min
Response: 2701318
Conc: 9.97 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 5.768 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#27 AR-1254-2

R.T.: 5.740 min
Delta R.T.: -0.010 min
Response: 2033653
Conc: 8.47 ng/ml




#28 AR-1254-3

R.T.: 6.169 min
Delta R.T.: -0.006 min
Response: 15709063
Conc: 21.81 ng/ml



#28 AR-1254-3


R.T.: 6.160 min
Delta R.T.: 0.006 min
Response: 1250783
Conc: 3.33 ng/ml



#29 AR-1254-4


R.T.: 6.434 min
Delta R.T.: 0.029 min
Response: 4503178
Conc: 10.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#29 AR-1254-4

R.T.: 6.366 min
Delta R.T.: -0.016 min
Response: 2311685
Conc: 10.62 ng/ml




#30 AR-1254-5

R.T.: 6.818 min
Delta R.T.: -0.008 min
Response: 10058519
Conc: 15.84 ng/ml



#30 AR-1254-5


R.T.: 6.789 min
Delta R.T.: -0.011 min
Response: 7162277
Conc: 22.19 ng/ml



#31 AR-1260-1


R.T.: 6.300 min
Delta R.T.: -0.009 min
Response: 19227446
Conc: 42.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#31 AR-1260-1

R.T.: 6.275 min
Delta R.T.: -0.011 min
Response: 6911716
Conc: 29.53 ng/ml




#32 AR-1260-2

R.T.: 6.489 min
Delta R.T.: -0.009 min
Response: 24656349
Conc: 44.37 ng/ml



#32 AR-1260-2


R.T.: 6.462 min
Delta R.T.: -0.011 min
Response: 11166798
Conc: 39.80 ng/ml



#33 AR-1260-3


R.T.: 6.858 min
Delta R.T.: -0.011 min
Response: 6397701
Conc: 13.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#33 AR-1260-3

R.T.: 6.617 min
Delta R.T.: -0.011 min
Response: 9538893
Conc: 36.19 ng/ml




#34 AR-1260-4

R.T.: 7.118 min
Delta R.T.: -0.011 min
Response: 3620925
Conc: 8.51 ng/ml



#34 AR-1260-4


R.T.: 7.088 min
Delta R.T.: -0.012 min
Response: 4277700
Conc: 20.11 ng/ml



#35 AR-1260-5


R.T.: 7.360 min
Delta R.T.: -0.010 min
Response: 9490797
Conc: 9.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#35 AR-1260-5

R.T.: 7.328 min
Delta R.T.: -0.012 min
Response: 7568688
Conc: 15.64 ng/ml




#36 AR-1262-1

R.T.: 6.858 min
Delta R.T.: -0.007 min
Response: 6397701
Conc: 9.72 ng/ml



#36 AR-1262-1


R.T.: 6.829 min
Delta R.T.: -0.011 min
Response: 5749460
Conc: 17.01 ng/ml



#37 AR-1262-2


R.T.: 7.360 min
Delta R.T.: -0.007 min
Response: 9490797
Conc: 8.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#37 AR-1262-2

R.T.: 7.328 min
Delta R.T.: -0.011 min
Response: 7568688
Conc: 13.24 ng/ml




#38 AR-1262-3

R.T.: 7.646 min
Delta R.T.: -0.008 min
Response: 1328687
Conc: 3.04 ng/ml



#38 AR-1262-3


R.T.: 7.611 min
Delta R.T.: -0.012 min
Response: 963998
Conc: 4.30 ng/ml



#39 AR-1262-4


R.T.: 7.709 min
Delta R.T.: -0.008 min
Response: 8103390
Conc: 10.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#39 AR-1262-4

R.T.: 7.674 min
Delta R.T.: -0.012 min
Response: 4144136
Conc: 10.12 ng/ml




#40 AR-1262-5

R.T.: 8.208 min
Delta R.T.: -0.008 min
Response: -2869190
Conc: N.D.



#40 AR-1262-5


R.T.: 8.168 min
Delta R.T.: -0.012 min
Response: 2473091
Conc: 13.38 ng/ml



#41 AR-1268-1


R.T.: 7.646 min
Delta R.T.: -0.008 min
Response: 1328687
Conc: 1.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#41 AR-1268-1

R.T.: 7.611 min
Delta R.T.: -0.012 min
Response: 963998
Conc: 1.49 ng/ml




#42 AR-1268-2

R.T.: 7.709 min
Delta R.T.: -0.009 min
Response: 8103390
Conc: 7.22 ng/ml



#42 AR-1268-2


R.T.: 7.674 min
Delta R.T.: -0.013 min
Response: 4144136
Conc: 7.13 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 7.929 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP




#43 AR-1268-3

R.T.: 7.884 min
Delta R.T.: -0.014 min
Response: 403923
Conc: 0.81 ng/ml




#44 AR-1268-4

R.T.: 8.208 min
Delta R.T.: -0.007 min
Response: -2869190
Conc: N.D.



#44 AR-1268-4


R.T.: 8.168 min
Delta R.T.: -0.012 min
Response: 2473091
Conc: 12.18 ng/ml



#45 AR-1268-5

R.T.: 8.509 min
Delta R.T.: -0.009 min
Response: 2255232
Conc: 0.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
STORMWATER-AQ-COMP



#45 AR-1268-5

R.T.: 8.464 min
Delta R.T.: -0.014 min
Response: 1046832
Conc: 0.71 ng/ml