

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0122118.M
 Title : GC EXTRACTABLES
 Last Update : Mon Dec 24 02:43:03 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0052870.D 1000 =P0052868.D 750 =P0052869.D
 250 =P0052871.D 50 =P0052872.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.306	1.395	1.371	1.367	1.441	1.376	E6	3.57
2) SA	Decachlorobiphenyl	7.047	6.762	6.663	7.045	7.537	7.011	E5	4.85
3) L1	AR-1016-1	3.225	3.236	3.212	3.424	3.648	3.349	E4	5.62
4) L1	AR-1016-2	4.968	5.022	5.017	5.378	5.861	5.249	E4	7.22
5) L1	AR-1016-3	2.976	2.977	2.973	3.236	3.502	3.133	E4	7.51
6) L1	AR-1016-4	2.404	2.400	2.402	2.517	2.547	2.454	E4	2.93
7) L1	AR-1016-5	2.302	2.296	2.299	2.436	2.468	2.360	E4	3.58
8) L2	AR-1221-1	1.350					1.350	E4	0.00
9) L2	AR-1221-2	9.747					9.747	E3	0.00
10) L2	AR-1221-3	3.035					3.035	E4	0.00
11) L3	AR-1232-1	2.352					2.352	E4	0.00
12) L3	AR-1232-2	1.205					1.205	E4	0.00
13) L3	AR-1232-3	2.277					2.277	E4	0.00
14) L3	AR-1232-4	1.045					1.045	E4	0.00
15) L3	AR-1232-5	6.986					6.986	E3	0.00
16) L4	AR-1242-1	2.739					2.739	E4	0.00
17) L4	AR-1242-2	4.219					4.219	E4	0.00
18) L4	AR-1242-3	2.515					2.515	E4	0.00
19) L4	AR-1242-4	2.012					2.012	E4	0.00
20) L4	AR-1242-5	1.929					1.929	E4	0.00
21) L5	AR-1248-1	2.038					2.038	E4	0.00
22) L5	AR-1248-2	2.858					2.858	E4	0.00
23) L5	AR-1248-3	3.223					3.223	E4	0.00
24) L5	AR-1248-4	3.317					3.317	E4	0.00
25) L5	AR-1248-5	3.160					3.160	E4	0.00
26) L6	AR-1254-1	3.841					3.841	E4	0.00
27) L6	AR-1254-2	5.682					5.682	E4	0.00
28) L6	AR-1254-3	5.971					5.971	E4	0.00
29) L6	AR-1254-4	4.072					4.072	E4	0.00
30) L6	AR-1254-5	4.152					4.152	E4	0.00
31) L7	AR-1260-1	4.166	3.861	3.868	4.135	4.299	4.066	E4	4.77
32) L7	AR-1260-2	4.667	4.446	4.416	4.902	5.148	4.716	E4	6.59
33) L7	AR-1260-3	3.145	2.972	2.981	3.171	3.551	3.164	E4	7.42
34) L7	AR-1260-4	3.405	3.293	3.236	3.481	3.607	3.404	E4	4.35
35) L7	AR-1260-5	6.627	6.337	6.191	6.891	6.831	6.575	E4	4.64
36) L8	AR-1262-1	5.480					5.480	E4	0.00
37) L8	AR-1262-2	8.511					8.511	E4	0.00
38) L8	AR-1262-3	5.832					5.832	E4	0.00
39) L8	AR-1262-4	2.725					2.725	E4	0.00
40) L8	AR-1262-5	2.987					2.987	E4	0.00
41) L9	AR-1268-1	1.266					1.266	E5	0.00
42) L9	AR-1268-2	1.191					1.191	E5	0.00
43) L9	AR-1268-3	1.044					1.044	E5	0.00
44) L9	AR-1268-4	3.912					3.912	E4	0.00
45) L9	AR-1268-5	3.204					3.204	E5	0.00

Signal #2 Calibration Files

500 =P0052870.D 1000 =P0052868.D 750 =P0052869.D
 250 =P0052871.D 50 =P0052872.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.850	2.901	2.850	2.931	3.265	2.959	E5	5.90
2) SA	Decachlorobiphenyl	2.721	2.577	2.585	2.874	3.316	2.814	E5	10.84
3) L1	AR-1016-1	0.774	0.743	0.750	0.840	1.018	0.825	E4	13.88
4) L1	AR-1016-2	1.494	1.480	1.384	1.615	1.820	1.559	E4	10.75
5) L1	AR-1016-3	6.858	6.847	6.905	7.305	7.562	7.095	E3	4.55

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122118.M
 Title : GC EXTRACTABLES
 Last Update : Mon Dec 24 02:43:03 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0052870.D 1000 =P0052868.D 750 =P0052869.D
 250 =P0052871.D 50 =P0052872.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	5.861	5.816	5.853	6.284	5.951	5.953	E3	3.22
7)	L1 AR-1016-5	8.000	7.279	7.261	5.915	8.194	7.330	E3	12.21
8)	L2 AR-1221-1	3.140					3.140	E3	0.00
9)	L2 AR-1221-2	2.343					2.343	E3	0.00
10)	L2 AR-1221-3	7.350					7.350	E3	0.00
11)	L3 AR-1232-1	5.563					5.563	E3	0.00
12)	L3 AR-1232-2	6.985					6.985	E3	0.00
13)	L3 AR-1232-3	3.088					3.088	E3	0.00
14)	L3 AR-1232-4	2.832					2.832	E3	0.00
15)	L3 AR-1232-5	3.684					3.684	E3	0.00
16)	L4 AR-1242-1	6.444					6.444	E3	0.00
17)	L4 AR-1242-2	1.279					1.279	E4	0.00
18)	L4 AR-1242-3	5.922					5.922	E3	0.00
19)	L4 AR-1242-4	5.897					5.897	E3	0.00
20)	L4 AR-1242-5	7.334					7.334	E3	0.00
21)	L5 AR-1248-1	5.326					5.326	E3	0.00
22)	L5 AR-1248-2	8.425					8.425	E3	0.00
23)	L5 AR-1248-3	8.533					8.533	E3	0.00
24)	L5 AR-1248-4	1.104					1.104	E4	0.00
25)	L5 AR-1248-5	1.013					1.013	E4	0.00
26)	L6 AR-1254-1	1.710					1.710	E4	0.00
27)	L6 AR-1254-2	1.475					1.475	E4	0.00
28)	L6 AR-1254-3	2.295					2.295	E4	0.00
29)	L6 AR-1254-4	1.299					1.299	E4	0.00
30)	L6 AR-1254-5	2.028					2.028	E4	0.00
31)	L7 AR-1260-1	1.515	1.418	1.434	1.630	1.889	1.577	E4	12.28
32)	L7 AR-1260-2	1.830	1.690	1.707	2.084	2.566	1.975	E4	18.51
33)	L7 AR-1260-3	1.745	1.637	1.640	1.890	2.067	1.796	E4	10.22
34)	L7 AR-1260-4	1.189	1.112	1.113	1.271	1.318	1.200	E4	7.74
35)	L7 AR-1260-5	2.779	2.602	2.593	2.946	3.140	2.812	E4	8.32
36)	L8 AR-1262-1	2.137					2.137	E4	0.00
37)	L8 AR-1262-2	3.449					3.449	E4	0.00
38)	L8 AR-1262-3	1.439					1.439	E4	0.00
39)	L8 AR-1262-4	2.683					2.683	E4	0.00
40)	L8 AR-1262-5	1.160					1.160	E4	0.00
41)	L9 AR-1268-1	4.783					4.783	E4	0.00
42)	L9 AR-1268-2	4.673					4.673	E4	0.00
43)	L9 AR-1268-3	4.025					4.025	E4	0.00
44)	L9 AR-1268-4	1.555					1.555	E4	0.00
45)	L9 AR-1268-5	1.160					1.160	E5	0.00

(#) = Out of Range