

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122118\
 Data File : P0052901.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Dec 2018 00:04
 Operator : SM/SJ
 Sample : J6516-01
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EXCAVATION-WATER

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 03:10:22 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 24 02:43:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.240	3.269	22088313	4700275	16.051	15.882
2)	SA Decachlor...	9.778	7.944	8242408	3423871	11.757	12.165

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122118\
Data File : PO052901.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Dec 2018 00:04
Operator : SM/SJ
Sample : J6516-01
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EXCAVATION-WATER

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 03:10:22 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 24 02:43:03 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.240 min
Delta R.T.: -0.002 min
Response: 22088313
Conc: 16.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXCAVATION-WATER




#1 Tetrachloro-m-xylene

R.T.: 3.269 min
Delta R.T.: 0.000 min
Response: 4700275
Conc: 15.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.778 min
Delta R.T.: -0.005 min
Response: 8242408
Conc: 11.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.944 min
Delta R.T.: -0.003 min
Response: 3423871
Conc: 12.17 ng/ml