

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD122222CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Dec 22 04:10:58 2022
 Response Via : Initial Calibration

Calibration Files

3 =PD073418.D 1 =PD073414.D 2 =PD073416.D
 4 =PD073420.D 5 =PD073422.D 6 =PD073424.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	2.074	2.223	2.174	1.999	1.978	1.968	2.070 E6	5.21
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin	3.208	3.202	3.303	3.192	3.232	3.295	3.239 E6	1.50
6) B beta-BHC	1.494	1.665	1.616	1.401	1.355	1.340	1.478 E6	9.28
7) B delta-BHC	3.402	3.228	3.376	3.452	3.568	3.709	3.456 E6	4.81
8) B Heptachlor ep...	2.956	3.092	3.047	2.894	2.901	2.930	2.970 E6	2.74
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane	2.912	3.040	3.034	2.863	2.903	2.988	2.957 E6	2.51
11) B cis-Chlordane	2.975	3.122	3.146	2.894	2.907	2.962	3.001 E6	3.60
12) B 4,4'-DDE	2.910	2.802	2.971	2.917	2.970	3.034	2.934 E6	2.68
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II	2.671	2.749	2.733	2.604	2.580	2.593	2.655 E6	2.78
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde	2.153	2.282	2.247	2.063	2.014	1.992	2.125 E6	5.75
19) B Endosulfan Su...	2.518	2.625	2.596	2.442	2.417	2.396	2.499 E6	3.85
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone	3.012	3.173	3.115	2.927	2.874	2.832	2.989 E6	4.55
22) Toxaphene-1							0.000	-1.00
23) Toxaphene-2							0.000	-1.00
24) Toxaphene-3							0.000	-1.00
25) Toxaphene-4							0.000	-1.00
26) Toxaphene-5							0.000	-1.00
27) SA Decachlorobip...	2.605	2.933	2.798	2.445	2.353	2.352	2.581 E6	9.40

Signal #2 Calibration Files

3 =PD073418.D 1 =PD073414.D 2 =PD073416.D
 4 =PD073420.D 5 =PD073422.D 6 =PD073424.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	7.578	7.915	8.243	6.966	7.178	7.572	7.575 E7	6.17
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin	5.257	5.960	5.698	4.711	4.504	4.426	5.093 E7	12.67
6) B beta-BHC	3.277	3.724	3.661	2.883	2.666	2.542	3.126 E7	16.18
7) B delta-BHC	6.960	8.312	7.757	6.261	6.164	6.700	7.026 E7	12.13
8) B Heptachlor ep...	4.722	5.130	5.108	4.129	3.930	3.832	4.475 E7	13.11
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane	4.343	4.726	4.663	3.973	3.847	3.877	4.238 E7	9.34
11) B cis-Chlordane	4.355	4.867	4.710	3.984	3.837	3.841	4.266 E7	10.54
12) B 4,4'-DDE	4.151	4.503	4.376	3.873	3.868	4.120	4.148 E7	6.22
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II	3.363	3.643	3.587	3.134	2.998	2.984	3.285 E7	8.84
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde	2.656	2.939	2.839	2.451	2.313	2.228	2.571 E7	11.18
19) B Endosulfan Su...	2.894	3.153	3.089	2.313	2.601	2.568	2.769 E7	11.89
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone	3.064	3.338	3.240	2.817	2.720	2.617	2.966 E7	9.86
22) Toxaphene-1							0.000	-1.00
23) Toxaphene-2							0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD122222CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Dec 22 04:10:58 2022
 Response Via : Initial Calibration

Calibration Files

3 =PD073418.D 1 =PD073414.D 2 =PD073416.D
 4 =PD073420.D 5 =PD073422.D 6 =PD073424.D

	Compound	3	1	2	4	5	6	Avg	%RSD
24)	Toxaphene-3							0.000	-1.00
25)	Toxaphene-4							0.000	-1.00
26)	Toxaphene-5							0.000	-1.00
27) SA	Decachlorobip...	2.041	2.221	2.169	1.905	1.797	1.806	1.990 E7	9.16

(#) = Out of Range