

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD122222CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Dec 22 04:14:38 2022
 Response Via : Initial Calibration

Calibration Files

3 =PD073417.D 1 =PD073413.D 2 =PD073415.D
 4 =PD073419.D 5 =PD073421.D 6 =PD073423.D

	Compound	3	1	2	4	5	6	Avg		%RSD
1) SA	Tetrachloro-m...	2.165	2.247	2.147	2.037	2.097	2.078	2.128	E6	3.49
2) A	alpha-BHC	3.378	3.070	3.142	3.386	3.666	3.756	3.400	E6	8.04
3) MA	gamma-BHC (Li...	3.147	2.926	2.998	3.160	3.389	3.462	3.180	E6	6.62
4) MA	Heptachlor	3.296	3.188	3.230	3.274	3.422	3.414	3.304	E6	2.89
5) MB	Aldrin	3.208	3.202	3.303	3.192	3.232	3.295	3.239	E6	1.50
6) B	beta-BHC	1.494	1.665	1.616	1.401	1.355	1.340	1.478	E6	9.28
7) B	delta-BHC	3.402	3.228	3.376	3.452	3.568	3.709	3.456	E6	4.81
8) B	Heptachlor ep...	2.956	3.092	3.047	2.894	2.901	2.930	2.970	E6	2.74
9) A	Endosulfan I	2.775	2.858	2.798	2.722	2.797	2.769	2.787	E6	1.60
10) B	trans-Chlordane	2.912	3.040	3.034	2.863	2.903	2.988	2.957	E6	2.51
11) B	cis-Chlordane	2.975	3.122	3.146	2.894	2.907	2.962	3.001	E6	3.60
12) B	4,4'-DDE	2.910	2.802	2.971	2.917	2.970	3.034	2.934	E6	2.68
13) MA	Dieldrin	3.050	2.923	2.958	3.080	3.201	3.143	3.059	E6	3.47
14) MA	Endrin	2.676	2.593	2.592	2.681	2.692	2.652	2.648	E6	1.69
15) B	Endosulfan II	2.671	2.749	2.733	2.604	2.580	2.593	2.655	E6	2.78
16) A	4,4'-DDD	2.424	2.222	2.335	2.405	2.513	2.475	2.396	E6	4.36
17) MA	4,4'-DDT	2.428	2.184	2.273	2.480	2.581	2.595	2.424	E6	6.83
18) B	Endrin aldehyde	2.153	2.282	2.247	2.063	2.014	1.992	2.125	E6	5.75
19) B	Endosulfan Su...	2.518	2.625	2.596	2.442	2.417	2.396	2.499	E6	3.85
20) A	Methoxychlor	1.376	1.398	1.379	1.341	1.288	1.226	1.335	E6	4.94
21) B	Endrin ketone	3.012	3.173	3.115	2.927	2.874	2.832	2.989	E6	4.55
22)	Toxaphene-1	1.344	1.373	1.282	1.345	1.312	1.282	1.323	E4	2.80
23)	Toxaphene-2	1.690	1.603	1.556	1.595	1.636	1.685	1.628	E4	3.26
24)	Toxaphene-3	2.132	2.062	2.101	2.240	2.086	2.032	2.109	E4	3.44
25)	Toxaphene-4	9.044	8.548	8.797	9.739	9.730	9.198	9.176	E4	5.30
26)	Toxaphene-5	5.345	4.913	5.004	5.597	5.652	5.825	5.389	E4	6.84
27) SA	Decachlorobip...	2.574	2.847	2.687	2.445	2.402	2.342	2.549	E6	7.52

Signal #2 Calibration Files

3 =PD073417.D 1 =PD073413.D 2 =PD073415.D
 4 =PD073419.D 5 =PD073421.D 6 =PD073423.D

	Compound	3	1	2	4	5	6	Avg		%RSD
1) SA	Tetrachloro-m...	7.909	8.033	7.691	6.333	7.084	7.839	7.481	E7	8.73
2) A	alpha-BHC	1.122	1.180	1.172	0.960	1.072	1.571	1.179	E8	17.63
3) MA	gamma-BHC (Li...	0.885	1.016	0.936	0.759	0.795	0.943	0.889	E8	10.90
4) MA	Heptachlor	6.145	6.992	6.507	5.347	5.255	5.094	5.890	E7	13.14
5) MB	Aldrin	5.257	5.960	5.698	4.711	4.504	4.426	5.093	E7	12.67
6) B	beta-BHC	3.277	3.724	3.661	2.883	2.666	2.542	3.126	E7	16.18
7) B	delta-BHC	6.960	8.312	7.757	6.261	6.164	6.700	7.026	E7	12.13
8) B	Heptachlor ep...	4.722	5.130	5.108	4.129	3.930	3.832	4.475	E7	13.11
9) A	Endosulfan I	4.115	4.517	4.333	3.704	3.638	3.512	3.970	E7	10.34
10) B	trans-Chlordane	4.343	4.726	4.663	3.973	3.847	3.877	4.238	E7	9.34
11) B	cis-Chlordane	4.355	4.867	4.710	3.984	3.837	3.841	4.266	E7	10.54
12) B	4,4'-DDE	4.151	4.503	4.376	3.873	3.868	4.120	4.148	E7	6.22
13) MA	Dieldrin	4.335	4.656	4.473	4.028	4.085	3.092	4.112	E7	13.43
14) MA	Endrin	3.494	3.879	3.609	3.036	3.121	3.045	3.364	E7	10.39
15) B	Endosulfan II	3.363	3.643	3.587	3.134	2.998	2.984	3.285	E7	8.84
16) A	4,4'-DDD	3.249	3.444	3.344	3.015	3.012	2.958	3.170	E7	6.40
17) MA	4,4'-DDT	3.089	3.173	3.107	2.934	2.887	2.916	3.018	E7	3.97
18) B	Endrin aldehyde	2.656	2.939	2.839	2.451	2.313	2.228	2.571	E7	11.18
19) B	Endosulfan Su...	2.894	3.153	3.089	2.313	2.601	2.568	2.769	E7	11.89
20) A	Methoxychlor	1.308	1.413	1.364	1.213	1.109	0.994	1.233	E7	12.97
21) B	Endrin ketone	3.064	3.338	3.240	2.817	2.720	2.617	2.966	E7	9.86
22)	Toxaphene-1	3.037	3.257	3.107	2.900	2.811	2.769	2.980	E5	6.29
23)	Toxaphene-2	2.008	2.462	2.049	1.881	1.738	1.705	1.974	E5	14.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD122222CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Dec 22 04:14:38 2022
 Response Via : Initial Calibration

Calibration Files

3 =PD073417.D 1 =PD073413.D 2 =PD073415.D
 4 =PD073419.D 5 =PD073421.D 6 =PD073423.D

	Compound	3	1	2	4	5	6	Avg		%RSD
24)	Toxaphene-3	6.331	6.553	6.377	6.218	6.232	6.388	6.350	E5	1.93
25)	Toxaphene-4	4.632	4.779	4.583	4.409	4.489	4.525	4.569	E5	2.81
26)	Toxaphene-5	3.752	4.050	3.944	3.856	3.921	3.809	3.889	E5	2.73
27) SA	Decachlorobip...	2.001	2.164	2.075	1.897	1.846	1.799	1.964	E7	7.17

(#) = Out of Range