

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122122.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 22 00:59:27 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0091584.D 1000 =P0091582.D 750 =P0091583.D
 250 =P0091585.D 50 =P0091586.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.004	2.895	2.924	2.968	2.681	2.894 E6	4.36
2) SA	Decachlorobiphenyl	2.128	1.973	2.041	2.212	2.129	2.096 E6	4.39
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.014	0.928	0.958	1.052	1.011	0.993 E5	4.95
27) L6	AR-1254-2	1.525	1.407	1.446	1.592	1.524	1.499 E5	4.85
28) L6	AR-1254-3	1.529	1.431	1.462	1.575	1.485	1.496 E5	3.79
29) L6	AR-1254-4	1.132	1.055	1.085	1.172	1.116	1.112 E5	4.04
30) L6	AR-1254-5	1.255	1.166	1.198	1.277	1.153	1.210 E5	4.51
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =P0091584.D 1000 =P0091582.D 750 =P0091583.D
 250 =P0091585.D 50 =P0091586.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.100	1.053	1.066	1.085	0.941	1.049 E6	6.02
2) SA	Decachlorobiphenyl	9.301	8.537	8.852	9.838	9.505	9.207 E5	5.62
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122122.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 22 00:59:27 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0091584.D 1000 =P0091582.D 750 =P0091583.D
 250 =P0091585.D 50 =P0091586.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	5.631	5.126	5.300	5.858	5.530	5.489 E4	5.20
27) L6	AR-1254-2	5.091	4.605	4.787	5.373	5.220	5.015 E4	6.27
28) L6	AR-1254-3	7.808	7.172	7.389	8.038	7.485	7.578 E4	4.54
29) L6	AR-1254-4	4.813	4.423	4.566	4.988	4.614	4.681 E4	4.73
30) L6	AR-1254-5	7.173	6.605	6.825	7.357	7.064	7.005 E4	4.21
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range