

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122122.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 22 02:06:03 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0091590.D 1000 =P0091588.D 750 =P0091589.D
 250 =P0091591.D 50 =P0091592.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.102	2.970	3.042	3.099	2.769	2.996 E6	4.60
2) SA	Decachlorobiphenyl	3.717	3.438	3.554	3.857	3.691	3.651 E6	4.40
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	3.045	2.844	2.931	3.113	2.995	2.986 E5	3.47
42) L9	AR-1268-2	2.747	2.571	2.641	2.803	2.686	2.690 E5	3.36
43) L9	AR-1268-3	2.351	2.182	2.254	2.402	2.261	2.290 E5	3.79
44) L9	AR-1268-4	1.057	0.966	1.000	1.087	1.043	1.031 E5	4.64
45) L9	AR-1268-5	7.590	7.229	7.369	7.638	7.072	7.380 E5	3.24

Signal #2 Calibration Files

500 =P0091590.D 1000 =P0091588.D 750 =P0091589.D
 250 =P0091591.D 50 =P0091592.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.126	1.075	1.100	1.124	0.966	1.078 E6	6.13
2) SA	Decachlorobiphenyl	1.600	1.474	1.524	1.671	1.671	1.588 E6	5.55
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122122.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 22 02:06:03 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0091590.D 1000 =P0091588.D 750 =P0091589.D
 250 =P0091591.D 50 =P0091592.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.423	1.325	1.367	1.471	1.411	1.399 E5	3.97
42) L9	AR-1268-2	1.289	1.214	1.241	1.325	1.254	1.265 E5	3.42
43) L9	AR-1268-3	1.102	1.022	1.051	1.143	1.106	1.085 E5	4.43
44) L9	AR-1268-4	4.684	4.320	4.454	4.896	4.399	4.551 E4	5.19
45) L9	AR-1268-5	3.307	3.130	3.191	3.367	3.259	3.251 E5	2.87

(#) = Out of Range