

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122123\
 Data File : P0100652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2023 21:53
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 22:37:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.350	3.646	61096864	39687290	19.581	19.422
2)	SA Decachlor...	10.090	8.706	41596547	31073903	19.911	20.588
Target Compounds							
3)	L1 AR-1016-1	5.519	4.748	200485	452043	2.336	7.305 #
4)	L1 AR-1016-2	5.537	4.748	150848	452043	1.182	5.127 #
5)	L1 AR-1016-3	5.593	4.937	648224	40762	8.219	0.873 #
6)	L1 AR-1016-4	5.701	4.985	51166	456098	0.840	11.114 #
7)	L1 AR-1016-5	5.998	5.202	-3557	619498	N.D.	11.926 #
8)	L2 AR-1221-1	4.536	3.875	257926	193184	6.977	7.693
9)	L2 AR-1221-2	4.656	3.949	1072485	781144	38.618	40.679
10)	L2 AR-1221-3	4.720	4.028	185350	48578	2.221	0.825 #
11)	L3 AR-1232-1	4.720	4.028	185350	48578	2.780	1.027 #
12)	L3 AR-1232-2	5.251	4.748	61816	452043	1.763	10.838 #
13)	L3 AR-1232-3	5.537	4.937	150848	40762	2.528	1.900
14)	L3 AR-1232-4	5.701	5.031	51166	243448	1.866	11.458 #
15)	L3 AR-1232-5	5.792	5.202	158911	619498	6.609	28.348 #
16)	L4 AR-1242-1	5.519	4.748	200485	452043	2.843	8.655 #
17)	L4 AR-1242-2	5.537	4.748	150848	452043	1.437	6.237 #
18)	L4 AR-1242-3	5.593	4.937	648224	40762	9.972	1.064 #
19)	L4 AR-1242-4	5.701	5.031	51166	243448	1.022	5.841 #
20)	L4 AR-1242-5	6.446	5.538	556230	161963	10.747	3.529 #
21)	L5 AR-1248-1	5.519	4.748	200485	452043	3.654	11.304 #
22)	L5 AR-1248-2	5.792	4.985	158911	456098	1.840	7.581 #
23)	L5 AR-1248-3	5.998	5.031	-3557	243448	N.D.	3.895 #
24)	L5 AR-1248-4	6.407	5.202	730547	619498	8.126	8.600
25)	L5 AR-1248-5	6.446	5.589	556230	439207	6.197	7.119
26)	L6 AR-1254-1	6.407	5.538	730547	161963	6.950	1.578 #
27)	L6 AR-1254-2	6.582	5.698	229547	350683	1.555	3.817 #
28)	L6 AR-1254-3	6.961	6.115	263063	95730	1.879	0.708 #
29)	L6 AR-1254-4	7.280	6.338	44059	137305	0.518	1.904 #
30)	L6 AR-1254-5	7.661	6.780	2893463	9516025	24.453	82.587 #
31)	L7 AR-1260-1	7.161f	6.239	2785761	213625	20.334	2.243 #
32)	L7 AR-1260-2	7.410	6.417	279252	483230	2.220	4.496 #
33)	L7 AR-1260-3	7.743	6.595	422497	239159	4.362	2.333 #
34)	L7 AR-1260-4	0.000	6.994f	0	31655	N.D.	0.388 #
35)	L7 AR-1260-5	8.328f	7.324	5883244	3275886	27.871	18.151 #
36)	L8 AR-1262-1	7.661	6.780	2893463	9516025	28.964	152.553 #
37)	L8 AR-1262-2	8.328f	6.994f	5883244	31655	24.792	0.280 #
38)	L8 AR-1262-3	8.598	7.593	642594	86415	3.957	1.122 #
39)	L8 AR-1262-4	8.692	7.660	107560	889979	0.834	6.024 #
40)	L8 AR-1262-5	9.342	8.145	197226	1130932	2.472	17.684 #
41)	L9 AR-1268-1	8.598	7.593	642594	86415	2.274	0.390 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122123\
 Data File : P0100652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2023 21:53
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 22:37:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.692	7.660	107560	889979	0.426	4.400 #
43) L9	AR-1268-3	8.960f	7.855	-319149	3431662	N.D.	18.189 #
44) L9	AR-1268-4	9.342	8.145	197226	1130932	2.519	16.844 #
45) L9	AR-1268-5	9.754	8.418	43682	2850526	0.061	5.494 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122123\
Data File : PO100652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2023 21:53
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 22:37:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.350 min
Delta R.T.: 0.001 min
Response: 61096864
Conc: 19.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.000 min
Response: 39687290
Conc: 19.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.090 min
Delta R.T.: -0.002 min
Response: 41596547
Conc: 19.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.706 min
Delta R.T.: 0.000 min
Response: 31073903
Conc: 20.59 ng/ml



#3 AR-1016-1


R.T.: 5.519 min
Delta R.T.: 0.001 min
Response: 200485
Conc: 2.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.748 min
Delta R.T.: 0.006 min
Response: 452043
Conc: 7.30 ng/ml




#4 AR-1016-2

R.T.: 5.537 min
Delta R.T.: -0.003 min
Response: 150848
Conc: 1.18 ng/ml



#4 AR-1016-2


R.T.: 4.748 min
Delta R.T.: -0.014 min
Response: 452043
Conc: 5.13 ng/ml



#5 AR-1016-3


R.T.: 5.593 min
Delta R.T.: -0.008 min
Response: 648224
Conc: 8.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.937 min
Delta R.T.: -0.003 min
Response: 40762
Conc: 0.87 ng/ml




#6 AR-1016-4

R.T.: 5.701 min
Delta R.T.: 0.001 min
Response: 51166
Conc: 0.84 ng/ml



#6 AR-1016-4


R.T.: 4.985 min
Delta R.T.: 0.004 min
Response: 456098
Conc: 11.11 ng/ml



#7 AR-1016-5


R.T.: 5.998 min
Delta R.T.: 0.002 min
Response: -3557
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.202 min
Delta R.T.: 0.004 min
Response: 619498
Conc: 11.93 ng/ml




#8 AR-1221-1

R.T.: 4.536 min
Delta R.T.: -0.022 min
Response: 257926
Conc: 6.98 ng/ml



#8 AR-1221-1


R.T.: 3.875 min
Delta R.T.: 0.014 min
Response: 193184
Conc: 7.69 ng/ml



#9 AR-1221-2


R.T.: 4.656 min
Delta R.T.: 0.011 min
Response: 1072485
Conc: 38.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.949 min
Delta R.T.: 0.001 min
Response: 781144
Conc: 40.68 ng/ml




#10 AR-1221-3

R.T.: 4.720 min
Delta R.T.: -0.001 min
Response: 185350
Conc: 2.22 ng/ml



#10 AR-1221-3


R.T.: 4.028 min
Delta R.T.: 0.003 min
Response: 48578
Conc: 0.83 ng/ml



#11 AR-1232-1


R.T.: 4.720 min
Delta R.T.: 0.000 min
Response: 185350
Conc: 2.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.028 min
Delta R.T.: 0.003 min
Response: 48578
Conc: 1.03 ng/ml




#12 AR-1232-2

R.T.: 5.251 min
Delta R.T.: 0.002 min
Response: 61816
Conc: 1.76 ng/ml



#12 AR-1232-2


R.T.: 4.748 min
Delta R.T.: -0.014 min
Response: 452043
Conc: 10.84 ng/ml



#13 AR-1232-3


R.T.: 5.537 min
Delta R.T.: -0.004 min
Response: 150848
Conc: 2.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.937 min
Delta R.T.: -0.003 min
Response: 40762
Conc: 1.90 ng/ml




#14 AR-1232-4

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 51166
Conc: 1.87 ng/ml



#14 AR-1232-4


R.T.: 5.031 min
Delta R.T.: 0.007 min
Response: 243448
Conc: 11.46 ng/ml



#15 AR-1232-5


R.T.: 5.792 min
Delta R.T.: -0.001 min
Response: 158911
Conc: 6.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.202 min
Delta R.T.: 0.004 min
Response: 619498
Conc: 28.35 ng/ml




#16 AR-1242-1

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 200485
Conc: 2.84 ng/ml



#16 AR-1242-1


R.T.: 4.748 min
Delta R.T.: 0.005 min
Response: 452043
Conc: 8.66 ng/ml



#17 AR-1242-2


R.T.: 5.537 min
Delta R.T.: -0.004 min
Response: 150848
Conc: 1.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.748 min
Delta R.T.: -0.014 min
Response: 452043
Conc: 6.24 ng/ml




#18 AR-1242-3

R.T.: 5.593 min
Delta R.T.: -0.010 min
Response: 648224
Conc: 9.97 ng/ml



#18 AR-1242-3


R.T.: 4.937 min
Delta R.T.: -0.004 min
Response: 40762
Conc: 1.06 ng/ml



#19 AR-1242-4


R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 51166
Conc: 1.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.031 min
Delta R.T.: 0.006 min
Response: 243448
Conc: 5.84 ng/ml




#20 AR-1242-5

R.T.: 6.446 min
Delta R.T.: 0.006 min
Response: 556230
Conc: 10.75 ng/ml



#20 AR-1242-5


R.T.: 5.538 min
Delta R.T.: -0.014 min
Response: 161963
Conc: 3.53 ng/ml



#21 AR-1248-1


R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 200485
Conc: 3.65 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.748 min
Delta R.T.: 0.006 min
Response: 452043
Conc: 11.30 ng/ml




#22 AR-1248-2

R.T.: 5.792 min
Delta R.T.: 0.000 min
Response: 158911
Conc: 1.84 ng/ml



#22 AR-1248-2


R.T.: 4.985 min
Delta R.T.: 0.004 min
Response: 456098
Conc: 7.58 ng/ml



#23 AR-1248-3


R.T.: 5.998 min
Delta R.T.: 0.002 min
Response: -3557
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.031 min
Delta R.T.: 0.008 min
Response: 243448
Conc: 3.89 ng/ml




#24 AR-1248-4

R.T.: 6.407 min
Delta R.T.: 0.006 min
Response: 730547
Conc: 8.13 ng/ml



#24 AR-1248-4


R.T.: 5.202 min
Delta R.T.: 0.005 min
Response: 619498
Conc: 8.60 ng/ml



#25 AR-1248-5


R.T.: 6.446 min
Delta R.T.: 0.007 min
Response: 556230
Conc: 6.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.589 min
Delta R.T.: -0.005 min
Response: 439207
Conc: 7.12 ng/ml




#26 AR-1254-1

R.T.: 6.407 min
Delta R.T.: 0.033 min
Response: 730547
Conc: 6.95 ng/ml



#26 AR-1254-1


R.T.: 5.538 min
Delta R.T.: -0.014 min
Response: 161963
Conc: 1.58 ng/ml



#27 AR-1254-2


R.T.: 6.582 min
Delta R.T.: -0.012 min
Response: 229547
Conc: 1.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.698 min
Delta R.T.: -0.003 min
Response: 350683
Conc: 3.82 ng/ml




#28 AR-1254-3

R.T.: 6.961 min
Delta R.T.: 0.000 min
Response: 263063
Conc: 1.88 ng/ml



#28 AR-1254-3


R.T.: 6.115 min
Delta R.T.: 0.006 min
Response: 95730
Conc: 0.71 ng/ml



#29 AR-1254-4


R.T.: 7.280 min
Delta R.T.: 0.033 min
Response: 44059
Conc: 0.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.338 min
Delta R.T.: -0.001 min
Response: 137305
Conc: 1.90 ng/ml




#30 AR-1254-5

R.T.: 7.661 min
Delta R.T.: -0.007 min
Response: 2893463
Conc: 24.45 ng/ml



#30 AR-1254-5


R.T.: 6.780 min
Delta R.T.: 0.020 min
Response: 9516025
Conc: 82.59 ng/ml



#31 AR-1260-1


R.T.: 7.161 min
Delta R.T.: 0.035 min
Response: 2785761
Conc: 20.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.239 min
Delta R.T.: 0.000 min
Response: 213625
Conc: 2.24 ng/ml




#32 AR-1260-2

R.T.: 7.410 min
Delta R.T.: 0.026 min
Response: 279252
Conc: 2.22 ng/ml



#32 AR-1260-2


R.T.: 6.417 min
Delta R.T.: -0.012 min
Response: 483230
Conc: 4.50 ng/ml



#33 AR-1260-3


R.T.: 7.743 min
Delta R.T.: -0.002 min
Response: 422497
Conc: 4.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.595 min
Delta R.T.: 0.011 min
Response: 239159
Conc: 2.33 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T.: 7.971 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 6.994 min
Delta R.T.: -0.065 min
Response: 31655
Conc: 0.39 ng/ml



#35 AR-1260-5


R.T.: 8.328 min
Delta R.T.: 0.039 min
Response: 5883244
Conc: 27.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.324 min
Delta R.T.: 0.023 min
Response: 3275886
Conc: 18.15 ng/ml




#36 AR-1262-1

R.T.: 7.661 min
Delta R.T.: -0.003 min
Response: 2893463
Conc: 28.96 ng/ml



#36 AR-1262-1


R.T.: 6.780 min
Delta R.T.: 0.021 min
Response: 9516025
Conc: 152.55 ng/ml



#37 AR-1262-2


R.T.: 8.328 min
Delta R.T.: 0.039 min
Response: 5883244
Conc: 24.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.994 min
Delta R.T.: -0.064 min
Response: 31655
Conc: 0.28 ng/ml




#38 AR-1262-3

R.T.: 8.598 min
Delta R.T.: -0.001 min
Response: 642594
Conc: 3.96 ng/ml



#38 AR-1262-3


R.T.: 7.593 min
Delta R.T.: 0.007 min
Response: 86415
Conc: 1.12 ng/ml



#39 AR-1262-4


R.T.: 8.692 min
Delta R.T.: 0.004 min
Response: 107560
Conc: 0.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.660 min
Delta R.T.: 0.010 min
Response: 889979
Conc: 6.02 ng/ml




#40 AR-1262-5

R.T.: 9.342 min
Delta R.T.: 0.003 min
Response: 197226
Conc: 2.47 ng/ml



#40 AR-1262-5


R.T.: 8.145 min
Delta R.T.: -0.005 min
Response: 1130932
Conc: 17.68 ng/ml



#41 AR-1268-1


R.T.: 8.598 min
Delta R.T.: 0.003 min
Response: 642594
Conc: 2.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.593 min
Delta R.T.: 0.008 min
Response: 86415
Conc: 0.39 ng/ml




#42 AR-1268-2

R.T.: 8.692 min
Delta R.T.: 0.002 min
Response: 107560
Conc: 0.43 ng/ml



#42 AR-1268-2


R.T.: 7.660 min
Delta R.T.: 0.009 min
Response: 889979
Conc: 4.40 ng/ml



#43 AR-1268-3


R.T.: 8.960 min
Delta R.T.: 0.044 min
Response: -319149
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.855 min
Delta R.T.: -0.004 min
Response: 3431662
Conc: 18.19 ng/ml




#44 AR-1268-4

R.T.: 9.342 min
Delta R.T.: 0.002 min
Response: 197226
Conc: 2.52 ng/ml



#44 AR-1268-4


R.T.: 8.145 min
Delta R.T.: -0.004 min
Response: 1130932
Conc: 16.84 ng/ml



#45 AR-1268-5

R.T.: 9.754 min
Delta R.T.: 0.000 min
Response: 43682
Conc: 0.06 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.418 min
Delta R.T.: -0.028 min
Response: 2850526
Conc: 5.49 ng/ml