

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122123\
 Data File : PO100636.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2023 17:09
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 17:45:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.350	3.646	163.2E6	106.0E6	52.307	51.897
2) SA Decachlor...	10.089	8.706	107.8E6	79343186	51.583	52.568
Target Compounds						
26) L6 AR-1254-1	6.375	5.552	48436320	52296756	460.778	509.567
27) L6 AR-1254-2	6.594	5.700	72551825	46478618	491.339	505.909
28) L6 AR-1254-3	6.960	6.108	68774459	68855754	491.148	509.145
29) L6 AR-1254-4	7.247	6.339	43813566	36326048	515.376	503.611
30) L6 AR-1254-5	7.667	6.759	55653803	57731199	470.341	501.036



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122123\
Data File : PO100636.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2023 17:09
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 17:45:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.350 min
Delta R.T.: 0.000 min
Response: 163212745
Conc: 52.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.000 min
Response: 106047049
Conc: 51.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.089 min
Delta R.T.: -0.002 min
Response: 107762887
Conc: 51.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.706 min
Delta R.T.: 0.000 min
Response: 79343186
Conc: 52.57 ng/ml



#26 AR-1254-1


R.T.: 6.375 min
Delta R.T.: 0.000 min
Response: 48436320
Conc: 460.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.552 min
Delta R.T.: 0.000 min
Response: 52296756
Conc: 509.57 ng/ml




#27 AR-1254-2

R.T.: 6.594 min
Delta R.T.: 0.000 min
Response: 72551825
Conc: 491.34 ng/ml



#27 AR-1254-2


R.T.: 5.700 min
Delta R.T.: -0.001 min
Response: 46478618
Conc: 505.91 ng/ml



#28 AR-1254-3


R.T.: 6.960 min
Delta R.T.: 0.000 min
Response: 68774459
Conc: 491.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.108 min
Delta R.T.: -0.001 min
Response: 68855754
Conc: 509.14 ng/ml




#29 AR-1254-4

R.T.: 7.247 min
Delta R.T.: 0.000 min
Response: 43813566
Conc: 515.38 ng/ml



#29 AR-1254-4


R.T.: 6.339 min
Delta R.T.: 0.000 min
Response: 36326048
Conc: 503.61 ng/ml



#30 AR-1254-5

R.T.: 7.667 min
Delta R.T.: 0.000 min
Response: 55653803
Conc: 470.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.759 min
Delta R.T.: 0.000 min
Response: 57731199
Conc: 501.04 ng/ml