

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122220\
 Data File : P0074041.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Dec 2020 18:42
 Operator : DD\AJ
 Sample : L5193-05
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LLEXSITUTV-022-002-SO

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 23 08:09:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 11 03:58:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.936	3.994	2163569	1252151	23.664	23.345
2)	SA Decachlor...	10.959	9.257	1933731	1092001	23.399	23.338
Target Compounds							
8)	L2 AR-1221-1	5.191	0.000	103687	0	97.861	N.D. #
9)	L2 AR-1221-2	5.289	0.000	22194	0	28.255	N.D. #
26)	L6 AR-1254-1	7.166	0.000	15308	0	4.033	N.D. #
29)	L6 AR-1254-4	8.097	0.000	10783	0	2.055	N.D. #
30)	L6 AR-1254-5	8.516	0.000	32601	0	6.330	N.D. #
32)	L7 AR-1260-2	8.237	0.000	14857	0	2.645	N.D. #
33)	L7 AR-1260-3	8.588	0.000	4835	0	1.058	N.D. #
36)	L8 AR-1262-1	8.588	0.000	4835	0	0.646	N.D. #
37)	L8 AR-1262-2	9.119f	0.000	3982	0	0.310	N.D. #
45)	L9 AR-1268-5	10.631	0.000	4198	0	0.128	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122220\
Data File : PO074041.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Dec 2020 18:42
Operator : DD\AJ
Sample : L5193-05
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LLEXSITUTV-022-002-SO

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 23 08:09:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 11 03:58:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.936 min
Delta R.T.: -0.005 min
Response: 2163569
Conc: 23.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
LLEXSITUTV-022-002-SO




#1 Tetrachloro-m-xylene

R.T.: 3.994 min
Delta R.T.: -0.009 min
Response: 1252151
Conc: 23.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.959 min
Delta R.T.: -0.010 min
Response: 1933731
Conc: 23.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.257 min
Delta R.T.: -0.008 min
Response: 1092001
Conc: 23.34 ng/ml



#8 AR-1221-1


R.T.: 5.191 min
Delta R.T.: 0.000 min
Response: 103687
Conc: 97.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
LLEXSITUTV-022-002-SO




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.259 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.289 min
Delta R.T.: -0.001 min
Response: 22194
Conc: 28.25 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.357 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 7.166 min
Delta R.T.: -0.011 min
Response: 15308
Conc: 4.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
LLEXSITUTV-022-002-SO




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.118 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 8.097 min
Delta R.T.: 0.009 min
Response: 10783
Conc: 2.06 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.932 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 8.516 min
Delta R.T.: -0.001 min
Response: 32601
Conc: 6.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
LLEXSITUTV-022-002-SO




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.359 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 8.237 min
Delta R.T.: 0.013 min
Response: 14857
Conc: 2.64 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 7.028 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 8.588 min
Delta R.T.: -0.003 min
Response: 4835
Conc: 1.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
LLEXSITUTV-022-002-SO




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 7.182 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 8.588 min
Delta R.T.: -0.002 min
Response: 4835
Conc: 0.65 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 7.357 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 9.119 min
Delta R.T.: -0.029 min
Response: 3982
Conc: 0.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
LLEXSITUTV-022-002-SO




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.907 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.631 min
Delta R.T.: -0.001 min
Response: 4198
Conc: 0.13 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 9.028 min
Response: 0
Conc: N.D.