

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122223\
 Data File : PO100715.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Dec 2023 19:37
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 22:25:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.645	157.9E6	99701394	50.603	48.791
2)	SA Decachlor...	10.088	8.705	103.4E6	75572395	49.510	50.069
Target Compounds							
21)	L5 AR-1248-1	5.516	4.740	25995545	20156063	473.850	504.045
22)	L5 AR-1248-2	5.791	4.979	41150597	29932593	476.460	497.528
23)	L5 AR-1248-3	5.995	5.022	42247193	31528083	481.321	504.374
24)	L5 AR-1248-4	6.399	5.196	45360380	36708502	504.526	509.623
25)	L5 AR-1248-5	6.438	5.592	45534331	31219394	507.287	506.053



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122223\
Data File : PO100715.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2023 19:37
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 22:25:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: 0.000 min
Response: 157895323
Conc: 50.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: -0.001 min
Response: 99701394
Conc: 48.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.088 min
Delta R.T.: -0.004 min
Response: 103431696
Conc: 49.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.705 min
Delta R.T.: -0.001 min
Response: 75572395
Conc: 50.07 ng/ml



#21 AR-1248-1


R.T.: 5.516 min
Delta R.T.: -0.001 min
Response: 25995545
Conc: 473.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.740 min
Delta R.T.: -0.002 min
Response: 20156063
Conc: 504.05 ng/ml




#22 AR-1248-2

R.T.: 5.791 min
Delta R.T.: 0.000 min
Response: 41150597
Conc: 476.46 ng/ml



#22 AR-1248-2


R.T.: 4.979 min
Delta R.T.: -0.002 min
Response: 29932593
Conc: 497.53 ng/ml



#23 AR-1248-3


R.T.: 5.995 min
Delta R.T.: -0.001 min
Response: 42247193
Conc: 481.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.022 min
Delta R.T.: -0.002 min
Response: 31528083
Conc: 504.37 ng/ml




#24 AR-1248-4

R.T.: 6.399 min
Delta R.T.: -0.001 min
Response: 45360380
Conc: 504.53 ng/ml



#24 AR-1248-4

R.T.: 5.196 min
Delta R.T.: -0.002 min
Response: 36708502
Conc: 509.62 ng/ml



#25 AR-1248-5

R.T.: 6.438 min
Delta R.T.: 0.000 min
Response: 45534331
Conc: 507.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

#25 AR-1248-5

R.T.: 5.592 min
Delta R.T.: -0.002 min
Response: 31219394
Conc: 506.05 ng/ml